

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090123\
 Data File : VU055044.D
 Acq On : 01 Sep 2023 16:51
 Operator : MD/SY
 Sample : VU0901WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK023

Quant Time: Sep 01 23:47:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 01 23:46:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	199935	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	195304	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	98314	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	43414	4.633	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.600%
7) Chloroethane-d5	1.907	69	49696	4.582	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.600%
11) 1,1-Dichloroethene-d2	2.563	65	26220	4.911	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.200%
20) 2-Butanone-d5	4.624	46	127905	47.036	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.080%
24) Chloroform-d	5.061	84	119365	4.800	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
26) 1,2-Dichloroethane-d4	5.698	65	61068	4.909	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
32) Benzene-d6	5.724	84	236916	4.914	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
36) 1,2-Dichloropropane-d6	6.688	67	78871	5.144	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.800%
41) Toluene-d8	7.894	98	214928	4.840	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.800%
43) trans-1,3-Dichloroprop...	8.180	79	19258	4.561	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.200%
46) 2-Hexanone-d5	8.634	63	81164	47.345	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.700%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	51249	4.753	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	71619	4.991	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.800%
Target Compounds						Qvalue
16) Methylene chloride	3.042	84	5180	0.367	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090123\
Data File : VU055044.D
Acq On : 01 Sep 2023 16:51
Operator : MD/SY
Sample : VU0901WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK023

Quant Time: Sep 01 23:47:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 01 23:46:10 2023
Response via : Initial Calibration

