

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090123\
 Data File : VU055050.D
 Acq On : 01 Sep 2023 19:14
 Operator : MD/SY
 Sample : 04209-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 01 23:49:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 01 23:46:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	186176	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	183823	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	93319	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	45436	5.207	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.200%
7) Chloroethane-d5	1.907	69	51987	5.147	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.000%
11) 1,1-Dichloroethene-d2	2.563	65	26199	5.270	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.400%
20) 2-Butanone-d5	4.628	46	157673	62.269	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.540%
24) Chloroform-d	5.058	84	123185	5.320	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.400%
26) 1,2-Dichloroethane-d4	5.698	65	67463	5.824	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.400%
32) Benzene-d6	5.724	84	250240	5.515	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.200%
36) 1,2-Dichloropropane-d6	6.689	67	79839	5.532	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.600%
41) Toluene-d8	7.894	98	220887	5.285	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
43) trans-1,3-Dichloroprop...	8.177	79	20929	5.266	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	105.400%
46) 2-Hexanone-d5	8.631	63	95912	59.443	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.880%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	59286	5.841	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	116.800%
66) 1,2-Dichlorobenzene-d4	12.193	152	79963	5.871	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.400%
Target Compounds						Qvalue
16) Methylene chloride	3.042	84	6199	0.471	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090123\
Data File : VU055050.D
Acq On : 01 Sep 2023 19:14
Operator : MD/SY
Sample : 04209-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 01 23:49:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 01 23:46:10 2023
Response via : Initial Calibration

