

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090123\
 Data File : VU055060.D
 Acq On : 01 Sep 2023 23:33
 Operator : MD/SY
 Sample : VU0901WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK024

Quant Time: Sep 02 00:23:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 02 00:23:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	198347	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	196253	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	99167	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	44877	4.827	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.600%
7) Chloroethane-d5	1.911	69	52007	4.833	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.600%
11) 1,1-Dichloroethene-d2	2.563	65	26359	4.976	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.600%
20) 2-Butanone-d5	4.627	46	148312	54.978	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.960%
24) Chloroform-d	5.058	84	121891	4.941	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.800%
26) 1,2-Dichloroethane-d4	5.701	65	64369	5.216	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.724	84	247210	5.103	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
36) 1,2-Dichloropropane-d6	6.688	67	80619	5.232	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.600%
41) Toluene-d8	7.897	98	216536	4.852	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.000%
43) trans-1,3-Dichloroprop...	8.180	79	20365	4.800	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.000%
46) 2-Hexanone-d5	8.630	63	91584	53.165	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.340%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	58838	5.430	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	75258	5.199	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090123\
Data File : VU055060.D
Acq On : 01 Sep 2023 23:33
Operator : MD/SY
Sample : VU0901WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK024

Quant Time: Sep 02 00:23:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 02 00:23:04 2023
Response via : Initial Calibration

