

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090219\
 Data File : VU034170.D
 Acq On : 01 Sep 2019 15:42
 Operator : JC/SP
 Sample : K4648-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-5

Quant Time: Sep 02 09:08:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	439716	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	692389	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	678463	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	342711	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	276099	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
35) Dibromofluoromethane	4.86	113	232885	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
50) Toluene-d8	7.55	98	902171	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
62) 4-Bromofluorobenzene	10.29	95	318538	45.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.08%	
Target Compounds						
11) Tert butyl alcohol	2.81	59	35370	27.223	ug/l	99
16) Acetone	2.31	43	13255	4.611	ug/l	98
19) Methyl tert-butyl Ether	2.99	73	22271	1.544	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090219\
Data File : VU034170.D
Acq On : 01 Sep 2019 15:42
Operator : JC/SP
Sample : K4648-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-5

Quant Time: Sep 02 09:08:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

