

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090219\
 Data File : VU034148.D
 Acq On : 01 Sep 2019 02:09
 Operator : JC/SP
 Sample : K4526-30
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T9-12-13-F1-(1.9)

Quant Time: Sep 01 05:48:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	453564	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	723328	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	728078	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	352751	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	288300	49.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.22%	
35) Dibromofluoromethane	4.86	113	239362	47.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.42%	
50) Toluene-d8	7.55	98	950981	50.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.24%	
62) 4-Bromofluorobenzene	10.29	95	334741	45.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.62%	
Target Compounds						
16) Acetone	2.31	43	16445	5.546	ug/l	99
20) Methylene Chloride	2.69	84	14539	2.590	ug/l	96
43) Isopropyl Acetate	5.53	43	64690	5.550	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090219\
Data File : VU034148.D
Acq On : 01 Sep 2019 02:09
Operator : JC/SP
Sample : K4526-30
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T9-12-13-F1-(1.9)

Quant Time: Sep 01 05:48:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.96 min Scan# 1204

Delta R.T. 0.00 min

Lab File: VU034148.D

Acq: 01 Sep 2019 02:09

Instrument :

MSVOA_U

ClientSampleId :


T9-12-13-F1-(1.9)

Tot Ion:168 Resp: 453564

Ion Ratio Lower Upper

168 100

99 54.0 42.3 63.5



Abundance Ion 167.90 (167.60 to 168.60): VU034148.D (-1185) (-)

Ion 98.90 (98.60 to 99.60): VU034148.D (-1185) (-)

4.96

200000

150000



100000

50000

0

4.90 5.00 5.10

Time-->



#16

Acetone

Concen: 5.546 ug/l

RT: 2.31 min Scan# 380

Delta R.T. 0.00 min

Lab File: VU034148.D


Acq: 01 Sep 2019 02:09

Tgt Ion: 43 Resp: 16445

Ion Ratio Lower Upper

43 100

58 33.9 26.8 40.2



Abundance Ion 43.00 (42.70 to 43.70): VU034148.D (-358) (-)

Ion 58.00 (57.70 to 58.70): VU034148.D (-358) (-)

2.31

10000

8000

6000


4000


2000

0

2.25 2.30 2.35 2.40

Time-->









#20
Methylene Chloride
Concen: 2.590 ug/l
RT: 2.69 min Scan# 498
Delta R.T. 0.01 min
Lab File: VU034148.D
Acq: 01 Sep 2019 02:09




Instrument :
MSVOA_U
ClientSampleId :
T9-12-13-F1-(1.9)


Tot Ion: 84 Resp: 14539
Ion Ratio Lower Upper
84 100
49 120.4 100.2 150.4
51 39.5 30.8 46.2
86 61.8 52.5 78.7



#33
1,2-Dichloroethane-d4
Concen: 49.607 ug/l
RT: 5.29 min Scan# 1305
Delta R.T. -0.00 min
Lab File: VU034148.D
Acq: 01 Sep 2019 02:09

Tgt Ion: 65 Resp: 288300
Ion Ratio Lower Upper
65 100
67 54.4 0.0 110.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.86 min Scan# 1485

Delta R.T. -0.00 min

Lab File: VU034148.D

Acq: 01 Sep 2019 02:09

Instrument :

MSVOA_U

ClientSampleId :

T9-12-13-F1-(1.9)


Tot Ion:114 Resp: 723328

Ion Ratio Lower Upper

114 100

63 20.9 0.0 42.0




88 16.3 0.0 32.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VU0

Ion 87.90 (87.60 to 88.60): VU0



#35

Dibromofluoromethane

Concen: 47.711 ug/l

RT: 4.86 min Scan# 1172

Delta R.T. 0.00 min

Lab File: VU034148.D

Acq: 01 Sep 2019 02:09


Tgt Ion:113 Resp: 239362

Ion Ratio Lower Upper

113 100

111 101.4 81.7 122.5



192 21.2 17.1 25.7



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 5.550 ug/l

RT: 5.53 min Scan# 1382

Delta R.T. 0.01 min

Lab File: VU034148.D

Acq: 01 Sep 2019 02:09

Instrument :

MSVOA_U

ClientSampleId :

T9-12-13-F1-(1.9)


Tot Ion: 43 Resp: 64690

Ion Ratio Lower Upper

43 100

61 21.9 17.2 25.8




87 13.7 10.8 16.2



Abundance Ion 43.00 (42.70 to 43.70): VU034148.D (-1382) (-)

Ion 61.00 (60.70 to 61.70): VU034148.D (-1382) (-)

Ion 87.00 (86.70 to 87.70): VU034148.D (-1382) (-)



#50

Toluene-d8

Concen: 50.624 ug/l

RT: 7.55 min Scan# 2008

Delta R.T. 0.00 min

Lab File: VU034148.D


Acq: 01 Sep 2019 02:09

Tgt Ion: 98 Resp: 950981

Ion Ratio Lower Upper



98 100


100 64.6 50.3 75.5



Abundance Ion 98.00 (97.70 to 98.70): VU034148.D (-2008) (-)

Ion 100.00 (99.70 to 100.70): VU034148.D (-2008) (-)





#62

4-Bromofluorobenzene

Concen: 45.809 ug/l

RT: 10.29 min Scan# 2861

Delta R.T. 0.00 min

Lab File: VU034148.D

Acq: 01 Sep 2019 02:09

Instrument :

MSVOA_U

ClientSampleId :

T9-12-13-F1-(1.9)


Tot Ion: 95 Resp: 334741

Ion Ratio Lower Upper

95 100

174 84.5 0.0 168.8

176 80.1 0.0 161.8




Abundance Ion 94.90 (94.60 to 95.60): VU034148.D (-2848) (-)

Ion 173.80 (173.50 to 174.50): VU034148.D (-2848) (-)

Ion 175.80 (175.50 to 176.50): VU034148.D (-2848) (-)

Time-->




Abundance Ion 94.90 (94.60 to 95.60): VU034148.D (-2768) (-)

Ion 173.80 (173.50 to 174.50): VU034148.D (-2768) (-)

Ion 175.80 (175.50 to 176.50): VU034148.D (-2768) (-)

Time-->



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 9.07 min Scan# 2482

Delta R.T. 0.00 min

Lab File: VU034148.D

Acq: 01 Sep 2019 02:09


Tgt Ion: 117 Resp: 728078

Ion Ratio Lower Upper

117 100

82 56.4 45.1 67.7

119 31.8 25.4 38.0




Abundance Ion 116.90 (116.60 to 117.60): VU034148.D (-2468) (-)

Ion 82.00 (81.70 to 82.70): VU034148.D (-2468) (-)

Ion 118.90 (118.60 to 119.60): VU034148.D (-2468) (-)

Time-->




Abundance Ion 116.90 (116.60 to 117.60): VU034148.D (-2389) (-)

Ion 82.00 (81.70 to 82.70): VU034148.D (-2389) (-)

Ion 118.90 (118.60 to 119.60): VU034148.D (-2389) (-)

Time-->



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 11.47 min Scan# 3227

Delta R.T. -0.00 min

Lab File: VU034148.D

Acq: 01 Sep 2019 02:09

Instrument :

MSVOA_U

ClientSampleId :

T9-12-13-F1-(1.9)


Tot Ion:152 Resp: 352751

Ion Ratio Lower Upper

152 100

115 56.6 41.1 123.3

150 158.3 0.0 348.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

