

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090220\
 Data File : VU040015.D
 Acq On : 02 Sep 2020 10:25
 Operator : SY/MD
 Sample : VU0902WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK30

Quant Time: Sep 03 01:27:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 03 01:16:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	113797	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	113958	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	51277	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	28211	4.54	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.80%
7) Chloroethane-d5	1.92	69	25976	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.58	63	50300	3.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.00%
20) 2-Butanone-d5	4.65	46	132609	53.36	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.72%
24) Chloroform-d	5.08	84	67298	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	5.72	65	44182	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.75	84	127500	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
36) 1,2-Dichloropropane-d6	6.70	67	42640	4.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.20%
41) Toluene-d8	7.91	98	109697	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.20%
43) trans-1,3-Dichloropropene-	8.19	79	18090	4.49	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.80%
46) 2-Hexanone-d5	8.64	63	95616	50.11	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.22%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	38515	5.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	42341	5.10	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.00%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	2465	0.214 ug/L	# 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090220\
Data File : VU040015.D
Acq On : 02 Sep 2020 10:25
Operator : SY/MD
Sample : VU0902WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK30

Quant Time: Sep 03 01:27:08 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 03 01:16:12 2020
Response via : Initial Calibration

