

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090221\
 Data File : VU044571.D
 Acq On : 02 Sep 2021 22:38
 Operator : SY/MD
 Sample : M3596-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F3W17

Quant Time: Sep 03 02:47:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 03 02:43:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	100769	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	103990	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	51033	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	40279	4.942	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.800%
7) Chloroethane-d5	1.919	69	38817	5.480	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.600%
11) 1,1-Dichloroethene-d2	2.575	65	16985	4.899	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.000%
20) 2-Butanone-d5	4.642	46	111131	49.305	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.620%
24) Chloroform-d	5.070	84	72898	5.058	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.200%
26) 1,2-Dichloroethane-d4	5.710	65	40517	5.156	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.200%
32) Benzene-d6	5.735	84	148884	5.084	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
36) 1,2-Dichloropropane-d6	6.697	67	45707	4.879	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.600%
41) Toluene-d8	7.902	98	130992	5.045	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
43) trans-1,3-Dichloroprop...	8.185	79	15072	4.480	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.600%
46) 2-Hexanone-d5	8.645	63	94082	51.713	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.420%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	40125	5.086	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.800%
66) 1,2-Dichlorobenzene-d4	12.198	152	50645	5.874	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.400%
Target Compounds						Qvalue
14) Carbon disulfide	2.800	76	2782	0.111	ug/L	98
47) Tetrachloroethene	8.558	164	5125	0.850	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090221\
Data File : VU044571.D
Acq On : 02 Sep 2021 22:38
Operator : SY/MD
Sample : M3596-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F3W17

Quant Time: Sep 03 02:47:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 03 02:43:14 2021
Response via : Initial Calibration

