

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090319\
 Data File : VU034192.D
 Acq On : 02 Sep 2019 11:20
 Operator : JC/SP
 Sample : K4544-25
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T16-17-B4-(0)

Manual Integrations
 APPROVED

MMDadoda
 9/3/2019 1:09:38 PM

Quant Time: Sep 03 06:59:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	431181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	693762	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	682695	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	317947	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	272806	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
35) Dibromofluoromethane	4.86	113	228636	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
50) Toluene-d8	7.55	98	909275	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	10.29	95	309081	44.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.20%	
Target Compounds						
20) Methylene Chloride	2.69	84	13262	2.485	ug/l	95
43) Isopropyl Acetate	5.53	43	70051	6.266	ug/l	99
95) Naphthalene	13.74	128	8504m	2.203	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090319\
Data File : VU034192.D
Acq On : 02 Sep 2019 11:20
Operator : JC/SP
Sample : K4544-25
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T16-17-B4-(0)

Manual Integrations
APPROVED

MMDadoda
9/3/2019 1:09:38 PM

Quant Time: Sep 03 06:59:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

