

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090319\
 Data File : VU034204.D
 Acq On : 02 Sep 2019 15:59
 Operator : JC/SP
 Sample : K4544-38
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T16-17-C4-(2.25)

Quant Time: Sep 03 07:18:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	432521	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	716059	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	715022	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	346307	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	283007	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
35) Dibromofluoromethane	4.86	113	235542	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
50) Toluene-d8	7.55	98	934752	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
62) 4-Bromofluorobenzene	10.29	95	323768	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
Target Compounds						
20) Methylene Chloride	2.69	84	8706	1.626	ug/l	90
43) Isopropyl Acetate	5.53	43	83094	7.201	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090319\
Data File : VU034204.D
Acq On : 02 Sep 2019 15:59
Operator : JC/SP
Sample : K4544-38
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T16-17-C4-(2.25)

Quant Time: Sep 03 07:18:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

