

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090319\
 Data File : VU034187.D
 Acq On : 02 Sep 2019 09:13
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 03 06:00:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.503	0.512	-1.8	83	0.00
3 P	Chloromethane	0.749	0.607	19.0	77	0.00
4 C	Vinyl Chloride	0.669	0.638	4.6#	81	0.00
5 T	Bromomethane	0.351	0.331	5.7	85	0.01
6 T	Chloroethane	0.409	0.384	6.1	83	0.01
7 T	Trichlorofluoromethane	0.775	0.777	-0.3	86	0.00
8 T	Diethyl Ether	0.332	0.338	-1.8	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.491	0.508	-3.5	89	0.00
10 T	Methyl Iodide	0.697	0.423	39.3#	50	0.00
11 T	Tert butyl alcohol	0.148	0.130	12.2	77	-0.01
12 CM	1,1-Dichloroethene	0.501	0.493	1.6#	85	0.00
13 T	Acrolein	0.062	0.038	38.7#	55	0.00
14 T	Allyl chloride	0.941	0.908	3.5	83	0.00
15 T	Acrylonitrile	0.339	0.345	-1.8	85	0.00
16 T	Acetone	0.327	0.392	-19.9	104	0.00
17 T	Carbon Disulfide	1.584	1.390	12.2	77	0.00
18 T	Methyl Acetate	0.845	0.805	4.7	88	0.00
19 T	Methyl tert-butyl Ether	1.640	1.685	-2.7	86	0.00
20 T	Methylene Chloride	0.619	0.597	3.6	86	0.00
21 T	trans-1,2-Dichloroethene	0.543	0.533	1.8	84	0.00
22 T	Diisopropyl ether	1.725	1.791	-3.8	87	0.00
23 T	Vinyl Acetate	1.481	1.520	-2.6	84	0.00
24 P	1,1-Dichloroethane	1.018	1.030	-1.2	86	0.00
25 T	2-Butanone	0.455	0.495	-8.8	88	0.00
26 T	2,2-Dichloropropane	0.828	0.833	-0.6	86	0.00
27 T	cis-1,2-Dichloroethene	0.610	0.602	1.3	85	0.00
28 T	Bromochloromethane	0.502	0.512	-2.0	88	0.00
29 T	Tetrahydrofuran	0.280	0.285	-1.8	83	0.00
30 C	Chloroform	1.016	0.995	2.1#	87	0.00
31 T	Cyclohexane	0.957	0.885	7.5	81	0.00
32 T	1,1,1-Trichloroethane	0.832	0.827	0.6	86	0.00
33 S	1,2-Dichloroethane-d4	0.641	0.625	2.5	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.347	0.356	-2.6	88	0.00
36 T	1,1-Dichloropropene	0.473	0.484	-2.3	84	0.00
37 T	Ethyl Acetate	0.530	0.545	-2.8	82	0.00
38 T	Carbon Tetrachloride	0.457	0.473	-3.5	85	0.00
39 T	Methylcyclohexane	0.557	0.577	-3.6	83	0.00
40 TM	Benzene	1.431	1.491	-4.2	85	0.00
41 T	Methacrylonitrile	0.272	0.293	-7.7	83	0.00
42 TM	1,2-Dichloroethane	0.494	0.514	-4.0	85	0.00
43 T	Isopropyl Acetate	0.806	0.830	-3.0	82	0.00
44 TM	Trichloroethene	0.376	0.396	-5.3	88	0.00
45 C	1,2-Dichloropropane	0.394	0.420	-6.6#	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090319\
 Data File : VU034187.D
 Acq On : 02 Sep 2019 09:13
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 03 06:00:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.254	0.270	-6.3	87	0.00
47 T	Bromodichloromethane	0.496	0.529	-6.7	87	0.00
48 T	Methyl methacrylate	0.389	0.403	-3.6	80	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	78	0.00
50 S	Toluene-d8	1.299	1.381	-6.3	86	0.00
51 T	4-Methyl-2-Pentanone	0.556	0.570	-2.5	82	0.00
52 CM	Toluene	0.881	0.951	-7.9#	87	0.00
53 T	t-1,3-Dichloropropene	0.551	0.577	-4.7	84	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.645	-7.1	86	0.00
55 T	1,1,2-Trichloroethane	0.373	0.390	-4.6	86	0.00
56 T	Ethyl methacrylate	0.557	0.582	-4.5	82	0.00
57 T	1,3-Dichloropropane	0.633	0.667	-5.4	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.206	0.196	4.9	67	0.00
59 T	2-Hexanone	0.420	0.458	-9.0	85	0.00
60 T	Dibromochloromethane	0.400	0.427	-6.7	88	0.00
61 T	1,2-Dibromoethane	0.393	0.407	-3.6	86	0.00
62 S	4-Bromofluorobenzene	0.505	0.524	-3.8	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.336	0.363	-8.0	89	0.00
65 PM	Chlorobenzene	1.029	1.069	-3.9	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.382	-2.4	86	0.00
67 C	Ethyl Benzene	1.744	1.849	-6.0#	85	0.00
68 T	m/p-Xylenes	0.663	0.727	-9.7	87	0.00
69 T	o-Xylene	0.636	0.678	-6.6	85	0.00
70 T	Styrene	1.091	1.229	-12.6	88	0.00
71 P	Bromoform	0.324	0.337	-4.0	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.358	3.377	-0.6	87	0.00
74 T	N-amyl acetate	1.413	1.401	0.8	82	0.00
75 P	1,1,2,2-Tetrachloroethane	1.285	1.221	5.0	87	0.00
76 T	1,2,3-Trichloropropane	1.127	1.018	9.7	79	0.00
77 T	Bromobenzene	0.903	0.880	2.5	88	0.00
78 T	n-propylbenzene	3.911	4.024	-2.9	89	0.00
79 T	2-Chlorotoluene	2.346	2.371	-1.1	88	0.00
80 T	1,3,5-Trimethylbenzene	2.847	2.949	-3.6	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.335	0.277	17.3	75	0.00
82 T	4-Chlorotoluene	2.698	2.777	-2.9	90	0.00
83 T	tert-Butylbenzene	2.802	2.782	0.7	86	0.00
84 T	1,2,4-Trimethylbenzene	2.827	2.981	-5.4	88	0.00
85 T	sec-Butylbenzene	3.359	3.426	-2.0	87	0.00
86 T	p-Isopropyltoluene	3.038	3.192	-5.1	88	0.00
87 T	1,3-Dichlorobenzene	1.599	1.644	-2.8	92	0.00
88 T	1,4-Dichlorobenzene	1.618	1.655	-2.3	93	0.00
89 T	n-Butylbenzene	2.678	2.923	-9.1	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090319\
Data File : VU034187.D
Acq On : 02 Sep 2019 09:13
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Sep 03 06:00:10 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.525	0.479	8.8	82	0.00
91 T	1,2-Dichlorobenzene	1.532	1.587	-3.6	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.256	0.252	1.6	85	0.00
93 T	1,2,4-Trichlorobenzene	0.937	1.118	-19.3	92	0.00
94 T	Hexachlorobutadiene	0.599	0.546	8.8	81	0.00
95 T	Naphthalene	2.584	3.001	-16.1	87	0.00
96 T	1,2,3-Trichlorobenzene	0.981	1.078	-9.9	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6