


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090319\
 Data File : VU034186.D
 Acq On : 02 Sep 2019 08:37
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Title : SW846 8260
 Last Update : Thu Aug 29 04:11:12 2019



AutoFind: Scans 2860, 2861, 2862; Background Corrected with Scan 2848

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	12599	PASS
75	95	30	60	48.4	36298	PASS
95	95	100	100	100.0	74946	PASS
96	95	5	9	6.9	5152	PASS
173	174	0.00	2	0.8	496	PASS
174	95	50	100	83.4	62506	PASS
175	174	5	9	7.3	4579	PASS
176	174	95	101	98.7	61701	PASS
177	176	5	9	6.8	4188	PASS