

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090320\
 Data File : VU040033.D
 Acq On : 03 Sep 2020 15:09
 Operator : SY/MD
 Sample : L3891-05MSD
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW289MSD

Quant Time: Sep 04 03:23:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 04 02:55:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	114273	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	111244	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	52192	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	27394	4.39	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.80%
7) Chloroethane-d5	1.92	69	25319	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.40%
11) 1,1-Dichloroethene-d2	2.58	63	72858	4.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.00%
20) 2-Butanone-d5	4.64	46	122335	49.02	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.04%
24) Chloroform-d	5.08	84	67959	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.72	65	42018	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
32) Benzene-d6	5.74	84	123707	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
36) 1,2-Dichloropropane-d6	6.70	67	42642	4.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.60%
41) Toluene-d8	7.91	98	105218	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
43) trans-1,3-Dichloropropene-	8.19	79	17446	4.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.80%
46) 2-Hexanone-d5	8.64	63	94531	50.75	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.50%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	38343	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	41587	4.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.40%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
12) 1,1-Dichloroethene	2.59	96	38960	4.406	ug/L	93
13) Acetone	2.64	43	2888	1.250	ug/L	97
33) Benzene	5.79	78	176290	4.430	ug/L	100
34) Trichloroethene	6.55	95	43889	4.111	ug/L	93
42) Toluene	7.98	91	179527	4.380	ug/L	99
51) Chlorobenzene	9.45	112	113639	4.466	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090320\
Data File : VU040033.D
Acq On : 03 Sep 2020 15:09
Operator : SY/MD
Sample : L3891-05MSD
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW289MSD

Quant Time: Sep 04 03:23:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 04 02:55:57 2020
Response via : Initial Calibration

