

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
 Data File : VU050616.D
 Acq On : 03 Sep 2022 16:46
 Operator : SY/MD
 Sample : N4481-03 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBVF3

Quant Time: Sep 05 01:08:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 05 01:05:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	281539	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	271168	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	97790	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	188684	78.356	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 156.720%#		
7) Chloroethane-d5	1.903	69	155165	95.287	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 190.580%#		
11) 1,1-Dichloroethene-d2	2.562	63	230186	54.093	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 108.180%		
21) 2-Butanone-d5	4.626	46	258566	146.763	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 146.760%#		
24) Chloroform-d	5.060	84	320106	70.139	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 140.280%#		
26) 1,2-Dichloroethane-d4	5.700	65	144102	51.670	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 103.340%		
32) Benzene-d6	5.726	84	456390	50.929	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 101.860%		
36) 1,2-Dichloropropane-d6	6.690	67	153178	52.063	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 104.120%		
41) Toluene-d8	7.899	98	366898	47.878	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 95.760%		
43) trans-1,3-Dichloroprop...	8.179	79	50693	41.642	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 83.280%		
47) 2-Hexanone-d5	8.636	63	66654	90.066	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 90.070%		
56) 1,1,2,2-Tetrachloroeth...	10.755	84	215089	48.310	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 96.620%		
66) 1,2-Dichlorobenzene-d4	12.192	152	119398	52.157	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 104.320%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
Data File : VU050616.D
Acq On : 03 Sep 2022 16:46
Operator : SY/MD
Sample : N4481-03 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBVF3

Quant Time: Sep 05 01:08:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 05 01:05:15 2022
Response via : Initial Calibration

