

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
 Data File : VU050624.D
 Acq On : 03 Sep 2022 19:55
 Operator : SY/MD
 Sample : N4363-16ME
 Misc : 5.56g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBVX0ME

Quant Time: Sep 05 01:09:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 05 01:05:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	453058	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	435880	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	158922	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	206329	53.245	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 106.500%		
7) Chloroethane-d5	1.893	69	85386	32.584	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	= 65.160%#		
11) 1,1-Dichloroethene-d2	2.536	63	270943	39.567	ug/L	-0.03
Spiked Amount	50.000	Range 60 - 125	Recovery	= 79.140%		
21) 2-Butanone-d5	4.642	46	412157	145.376	ug/L	0.01
Spiked Amount	100.000	Range 40 - 130	Recovery	= 145.380%#		
24) Chloroform-d	5.060	84	433771	59.062	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 118.120%		
26) 1,2-Dichloroethane-d4	5.700	65	280874	62.584	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 125.160%#		
32) Benzene-d6	5.719	84	893214	62.009	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 124.020%		
36) 1,2-Dichloropropane-d6	6.690	67	308557	65.244	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 130.480%#		
41) Toluene-d8	7.896	98	757967	61.533	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 123.060%#		
43) trans-1,3-Dichloroprop...	8.182	79	99668	50.935	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 101.860%		
47) 2-Hexanone-d5	8.658	63	240517	202.186	ug/L	0.02
Spiked Amount	100.000	Range 45 - 130	Recovery	= 202.190%#		
56) 1,1,2,2-Tetrachloroeth...	10.764	84	455554	63.654	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 127.300%#		
66) 1,2-Dichlorobenzene-d4	12.198	152	248143	66.701	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 133.400%#		
Target Compounds						Qvalue
46) Tetrachloroethene	8.552	164	6963	2.579	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
Data File : VU050624.D
Acq On : 03 Sep 2022 19:55
Operator : SY/MD
Sample : N4363-16ME
Misc : 5.56g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBVX0ME

Quant Time: Sep 05 01:09:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 05 01:05:15 2022
Response via : Initial Calibration

