

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
 Data File : VU050606.D
 Acq On : 03 Sep 2022 12:52
 Operator : SY/MD
 Sample : VU0903WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK180

Quant Time: Sep 05 01:06:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 05 01:05:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	281541	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	269215	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	107378	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	103231	42.869	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	85.740%	
7) Chloroethane-d5	1.906	69	78069	47.942	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	95.880%	
11) 1,1-Dichloroethene-d2	2.565	63	149232	35.069	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	70.140%	
21) 2-Butanone-d5	4.629	46	168254	95.501	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	95.500%	
24) Chloroform-d	5.063	84	217702	47.701	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.400%	
26) 1,2-Dichloroethane-d4	5.703	65	140263	50.293	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.580%	
32) Benzene-d6	5.726	84	444169	49.925	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.840%	
36) 1,2-Dichloropropane-d6	6.690	67	148163	50.724	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	101.440%	
41) Toluene-d8	7.899	98	376083	49.432	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	98.860%	
43) trans-1,3-Dichloroprop...	8.179	79	51827	42.883	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	85.760%	
47) 2-Hexanone-d5	8.639	63	66960	91.136	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	91.140%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	210000	47.509	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	95.020%	
66) 1,2-Dichlorobenzene-d4	12.192	152	126213	50.211	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	100.420%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
Data File : VU050606.D
Acq On : 03 Sep 2022 12:52
Operator : SY/MD
Sample : VU0903WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK180

Quant Time: Sep 05 01:06:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 05 01:05:15 2022
Response via : Initial Calibration

