

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
 Data File : VU050607.D
 Acq On : 03 Sep 2022 13:15
 Operator : SY/MD
 Sample : VU0903WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK181

Quant Time: Sep 05 01:06:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 05 01:05:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	263744	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	256982	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	96959	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	89503	39.676	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.360%
7) Chloroethane-d5	1.903	69	67017	43.932	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.860%
11) 1,1-Dichloroethene-d2	2.562	63	126189	31.655	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.320%
21) 2-Butanone-d5	4.629	46	179642	108.845	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.850%
24) Chloroform-d	5.064	84	218934	51.208	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.420%
26) 1,2-Dichloroethane-d4	5.703	65	143877	55.070	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.140%
32) Benzene-d6	5.726	84	438724	51.660	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.320%
36) 1,2-Dichloropropane-d6	6.690	67	152776	54.793	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.580%
41) Toluene-d8	7.896	98	351599	48.414	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.820%
43) trans-1,3-Dichloroprop...	8.179	79	53963	46.776	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.560%
47) 2-Hexanone-d5	8.636	63	72174	102.908	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.910%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	228374	54.125	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.260%
66) 1,2-Dichlorobenzene-d4	12.192	152	121929	53.719	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
Data File : VU050607.D
Acq On : 03 Sep 2022 13:15
Operator : SY/MD
Sample : VU0903WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK181

Quant Time: Sep 05 01:06:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 05 01:05:15 2022
Response via : Initial Calibration

