

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
 Data File : VU050614.D
 Acq On : 03 Sep 2022 16:00
 Operator : SY/MD
 Sample : N4481-01 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBVF1

Quant Time: Sep 05 01:08:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 05 01:05:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	317279	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	314253	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	114986	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	131875	48.596	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	97.200%	
7) Chloroethane-d5	1.903	69	92470	50.389	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	100.780%	
11) 1,1-Dichloroethene-d2	2.562	63	156346	32.602	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	65.200%	
21) 2-Butanone-d5	4.626	46	185887	93.625	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	93.630%	
24) Chloroform-d	5.060	84	241533	46.961	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.920%	
26) 1,2-Dichloroethane-d4	5.700	65	155110	49.352	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.700%	
32) Benzene-d6	5.726	84	500796	48.222	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.440%	
36) 1,2-Dichloropropane-d6	6.690	67	176623	51.801	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	103.600%	
41) Toluene-d8	7.896	98	427819	48.173	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	96.340%	
43) trans-1,3-Dichloroprop...	8.179	79	58684	41.597	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	83.200%	
47) 2-Hexanone-d5	8.636	63	83842	97.758	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	97.760%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	255806	49.578	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	99.160%	
66) 1,2-Dichlorobenzene-d4	12.192	152	139602	51.863	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	103.720%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
Data File : VU050614.D
Acq On : 03 Sep 2022 16:00
Operator : SY/MD
Sample : N4481-01 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBVF1

Quant Time: Sep 05 01:08:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 05 01:05:15 2022
Response via : Initial Calibration

