

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
 Data File : VU050623.D
 Acq On : 03 Sep 2022 19:31
 Operator : SY/MD
 Sample : N4362-02MERE
 Misc : 5.47g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0FB5MERE

Quant Time: Sep 05 01:09:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 05 01:05:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	408245	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	385095	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	144752	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	280763	80.407	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 160.820%#		
7) Chloroethane-d5	1.893	69	116174	49.200	ug/L	-0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	= 98.400%		
11) 1,1-Dichloroethene-d2	2.533	63	376785	61.063	ug/L	-0.04
Spiked Amount	50.000	Range 60 - 125	Recovery	= 122.120%		
21) 2-Butanone-d5	4.642	46	488151	191.081	ug/L	0.01
Spiked Amount	100.000	Range 40 - 130	Recovery	= 191.080%#		
24) Chloroform-d	5.060	84	541974	81.896	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 163.800%#		
26) 1,2-Dichloroethane-d4	5.700	65	342713	84.745	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 169.500%#		
32) Benzene-d6	5.719	84	1129207	88.731	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 177.460%#		
36) 1,2-Dichloropropane-d6	6.690	67	382456	91.535	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 183.080%#		
41) Toluene-d8	7.896	98	979814	90.033	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 180.060%#		
43) trans-1,3-Dichloroprop...	8.182	79	124690	72.126	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 144.260%#		
47) 2-Hexanone-d5	8.661	63	293307	279.078	ug/L	0.03
Spiked Amount	100.000	Range 45 - 130	Recovery	= 279.080%#		
56) 1,1,2,2-Tetrachloroeth...	10.764	84	550378	87.046	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 174.100%#		
66) 1,2-Dichlorobenzene-d4	12.195	152	315234	93.029	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 186.060%#		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090322\
Data File : VU050623.D
Acq On : 03 Sep 2022 19:31
Operator : SY/MD
Sample : N4362-02MERE
Misc : 5.47g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0FB5MERE

Quant Time: Sep 05 01:09:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 05 01:05:15 2022
Response via : Initial Calibration

