

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090324\
 Data File : VU060584.D
 Acq On : 03 Sep 2024 14:12
 Operator : MD/SY
 Sample : VU0903WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0903WBL01

Quant Time: Sep 04 05:42:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U080524DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Aug 06 11:07:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.113	96	30307	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.631	95	8990	0.871	ug/l	0.00
Spiked Amount 1.000			Recovery	=	87.000%	
68) 1,2-Dichlorobenzene-d4	12.193	152	8703	0.776	ug/l	0.00
Spiked Amount 1.000			Recovery	=	78.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090324\
Data File : VU060584.D
Acq On : 03 Sep 2024 14:12
Operator : MD/SY
Sample : VU0903WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0903WBL01

Quant Time: Sep 04 05:42:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U080524DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Aug 06 11:07:28 2024
Response via : Initial Calibration

