

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090324\
 Data File : VU060587.D
 Acq On : 03 Sep 2024 16:05
 Operator : MD/SY
 Sample : P3810-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-RVOA-WS

Quant Time: Sep 04 05:44:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U080524DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Aug 06 11:07:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.110	96	33773	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.631	95	12168	1.057	ug/l	0.00
Spiked Amount 1.000			Recovery	=	106.000%	
68) 1,2-Dichlorobenzene-d4	12.187	152	11604	0.928	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.000%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.599	62	646368	43.533	ug/l	100
9) 1,1-Dichloroethene	2.573	96	62518	6.192	ug/l	98
15) Methylene Chloride	3.039	84	131163	10.377	ug/l	97
16) trans-1,2-Dichloroethene	3.345	96	176523	16.317	ug/l	94
22) cis-1,2-Dichloroethene	4.660	96	88299	7.619	ug/l	90
28) 1,1,1-Trichloroethane	5.309	97	159855	9.455	ug/l	99
30) Carbon Tetrachloride	5.518	117	298601	21.672	ug/l	97
35) Benzene	5.766	78	288050	6.900	ug/l	99
36) 1,2-Dichloroethane	5.785	62	100167	8.635	ug/l	97
37) Trichloroethene	6.534	130	152882	14.794	ug/l	98
38) 1,2-Dichloropropane	6.782	63	59760	5.292	ug/l	98
49) Toluene	7.962	92	421070	18.192	ug/l	99
52) 1,1,2-Trichloroethane	8.393	97	149351	17.810	ug/l	98
58) Tetrachloroethene	8.547	164	48152	5.214	ug/l	98
59) Chlorobenzene	9.441	112	227177	9.065	ug/l	99
63) Ethyl Benzene	9.563	91	671877	13.709	ug/l	99
64) m/p-Xylenes	9.688	106	412700	21.229	ug/l	99
65) o-Xylene	10.094	106	153472	8.458	ug/l	97
66) Styrene	10.106	104	344273	10.911	ug/l	98
83) 1,4-Dichlorobenzene	11.830	146	225124	10.563	ug/l	99
85) 1,2-Dichlorobenzene	12.203	146	319395	15.380	ug/l	100
87) 1,2,4-Trichlorobenzene	13.833	180	82922	6.982	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090324\
Data File : VU060587.D
Acq On : 03 Sep 2024 16:05
Operator : MD/SY
Sample : P3810-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-RVOA-WS

Quant Time: Sep 04 05:44:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U080524DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Aug 06 11:07:28 2024
Response via : Initial Calibration

