

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090324\
 Data File : VU060591.D
 Acq On : 03 Sep 2024 17:42
 Operator : MD/SY
 Sample : P3810-01DL 2X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-RVOA-WSDL

Quant Time: Sep 04 05:46:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U080524DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Aug 06 11:07:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.110	96	29113	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.631	95	10592	1.068	ug/l	0.00
Spiked Amount 1.000			Recovery	=	107.000%	
68) 1,2-Dichlorobenzene-d4	12.187	152	10714	0.994	ug/l	0.00
Spiked Amount 1.000			Recovery	=	99.000%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.602	62	305409	23.862	ug/l	99
9) 1,1-Dichloroethene	2.576	96	29947	3.441	ug/l	99
15) Methylene Chloride	3.039	84	64119	5.885	ug/l	94
16) trans-1,2-Dichloroethene	3.348	96	83888	8.995	ug/l	94
22) cis-1,2-Dichloroethene	4.663	96	41964	4.200	ug/l	89
28) 1,1,1-Trichloroethane	5.309	97	76829	5.272	ug/l	99
30) Carbon Tetrachloride	5.518	117	146096	12.301	ug/l	97
35) Benzene	5.769	78	134666	3.742	ug/l	99
36) 1,2-Dichloroethane	5.788	62	47026	4.703	ug/l	95
37) Trichloroethene	6.534	130	76252	8.560	ug/l	98
38) 1,2-Dichloropropane	6.782	63	28039	2.880	ug/l	96
49) Toluene	7.962	92	195369	9.792	ug/l	98
52) 1,1,2-Trichloroethane	8.393	97	75408	10.432	ug/l	99
58) Tetrachloroethene	8.547	164	23785	2.988	ug/l	96
59) Chlorobenzene	9.441	112	109346	5.062	ug/l	100
63) Ethyl Benzene	9.563	91	301564	7.377	ug/l	100
64) m/p-Xylenes	9.689	106	190906	11.808	ug/l	100
65) o-Xylene	10.094	106	71490	4.774	ug/l	93
66) Styrene	10.107	104	159568	6.084	ug/l	98
83) 1,4-Dichlorobenzene	11.830	146	104529	5.689	ug/l	100
85) 1,2-Dichlorobenzene	12.206	146	152417	8.514	ug/l	99
87) 1,2,4-Trichlorobenzene	13.836	180	36640	3.579	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090324\
Data File : VU060591.D
Acq On : 03 Sep 2024 17:42
Operator : MD/SY
Sample : P3810-01DL 2X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-RVOA-WSDL

Quant Time: Sep 04 05:46:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U080524DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Aug 06 11:07:28 2024
Response via : Initial Calibration

