

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090419\
 Data File : VU034213.D
 Acq On : 03 Sep 2019 09:27
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 04 04:40:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.503	0.465	7.6	82	0.00
3 P	Chloromethane	0.749	0.567	24.3#	78	0.00
4 C	Vinyl Chloride	0.669	0.578	13.6#	80	0.00
5 T	Bromomethane	0.351	0.312	11.1	87	0.01
6 T	Chloroethane	0.409	0.354	13.4	84	0.01
7 T	Trichlorofluoromethane	0.775	0.716	7.6	86	0.00
8 T	Diethyl Ether	0.332	0.310	6.6	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.491	0.470	4.3	90	0.00
10 T	Methyl Iodide	0.697	0.454	34.9#	59	0.00
11 T	Tert butyl alcohol	0.148	0.104	29.7#	67	-0.01
12 CM	1,1-Dichloroethene	0.501	0.457	8.8#	86	0.00
13 T	Acrolein	0.062	0.035	43.5#	56	0.00
14 T	Allyl chloride	0.941	0.841	10.6	84	0.00
15 T	Acrylonitrile	0.339	0.285	15.9	77	0.00
16 T	Acetone	0.327	0.307	6.1	89	0.00
17 T	Carbon Disulfide	1.584	1.269	19.9	77	0.00
18 T	Methyl Acetate	0.845	0.682	19.3	81	0.00
19 T	Methyl tert-butyl Ether	1.640	1.559	4.9	86	0.00
20 T	Methylene Chloride	0.619	0.556	10.2	87	0.00
21 T	trans-1,2-Dichloroethene	0.543	0.496	8.7	85	0.00
22 T	Diisopropyl ether	1.725	1.681	2.6	89	0.00
23 T	Vinyl Acetate	1.481	1.356	8.4	81	0.00
24 P	1,1-Dichloroethane	1.018	0.955	6.2	87	0.00
25 T	2-Butanone	0.455	0.402	11.6	78	0.00
26 T	2,2-Dichloropropane	0.828	0.770	7.0	86	0.00
27 T	cis-1,2-Dichloroethene	0.610	0.562	7.9	86	0.00
28 T	Bromochloromethane	0.502	0.473	5.8	88	0.00
29 T	Tetrahydrofuran	0.280	0.234	16.4	74	0.00
30 C	Chloroform	1.016	0.934	8.1#	89	0.00
31 T	Cyclohexane	0.957	0.819	14.4	82	0.00
32 T	1,1,1-Trichloroethane	0.832	0.777	6.6	88	0.00
33 S	1,2-Dichloroethane-d4	0.641	0.609	5.0	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.347	0.350	-0.9	95	0.00
36 T	1,1-Dichloropropene	0.473	0.451	4.7	86	0.00
37 T	Ethyl Acetate	0.530	0.453	14.5	75	0.00
38 T	Carbon Tetrachloride	0.457	0.441	3.5	87	0.00
39 T	Methylcyclohexane	0.557	0.525	5.7	83	0.00
40 TM	Benzene	1.431	1.386	3.1	87	0.00
41 T	Methacrylonitrile	0.272	0.250	8.1	78	0.00
42 TM	1,2-Dichloroethane	0.494	0.469	5.1	85	0.00
43 T	Isopropyl Acetate	0.806	0.715	11.3	78	0.00
44 TM	Trichloroethene	0.376	0.370	1.6	91	0.00
45 C	1,2-Dichloropropane	0.394	0.389	1.3#	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090419\
 Data File : VU034213.D
 Acq On : 03 Sep 2019 09:27
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 04 04:40:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.254	0.242	4.7	86	0.00
47 T	Bromodichloromethane	0.496	0.496	0.0	90	0.00
48 T	Methyl methacrylate	0.389	0.345	11.3	75	0.00
49 T	1,4-Dioxane	0.010	0.007	30.0#	64	0.00
50 S	Toluene-d8	1.299	1.344	-3.5	93	0.00
51 T	4-Methyl-2-Pentanone	0.556	0.476	14.4	75	0.00
52 CM	Toluene	0.881	0.877	0.5#	88	0.00
53 T	t-1,3-Dichloropropene	0.551	0.522	5.3	84	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.589	2.2	86	0.00
55 T	1,1,2-Trichloroethane	0.373	0.362	2.9	88	0.00
56 T	Ethyl methacrylate	0.557	0.519	6.8	80	0.00
57 T	1,3-Dichloropropane	0.633	0.599	5.4	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.206	0.171	17.0	64	0.00
59 T	2-Hexanone	0.420	0.370	11.9	76	0.00
60 T	Dibromochloromethane	0.400	0.391	2.3	89	0.00
61 T	1,2-Dibromoethane	0.393	0.359	8.7	83	0.00
62 S	4-Bromofluorobenzene	0.505	0.501	0.8	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.336	0.349	-3.9	92	0.00
65 PM	Chlorobenzene	1.029	1.003	2.5	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.364	2.4	88	0.00
67 C	Ethyl Benzene	1.744	1.738	0.3#	86	0.00
68 T	m/p-Xylenes	0.663	0.682	-2.9	88	0.00
69 T	o-Xylene	0.636	0.642	-0.9	86	0.00
70 T	Styrene	1.091	1.147	-5.1	88	0.00
71 P	Bromoform	0.324	0.304	6.2	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.358	3.211	4.4	86	0.00
74 T	N-amyl acetate	1.413	1.259	10.9	77	0.00
75 P	1,1,2,2-Tetrachloroethane	1.285	1.096	14.7	82	0.00
76 T	1,2,3-Trichloropropane	1.127	0.944	16.2	77	0.00
77 T	Bromobenzene	0.903	0.838	7.2	87	0.00
78 T	n-propylbenzene	3.911	3.826	2.2	88	0.00
79 T	2-Chlorotoluene	2.346	2.227	5.1	87	0.00
80 T	1,3,5-Trimethylbenzene	2.847	2.809	1.3	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.335	0.297	11.3	84	0.00
82 T	4-Chlorotoluene	2.698	2.614	3.1	89	0.00
83 T	tert-Butylbenzene	2.802	2.646	5.6	85	0.00
84 T	1,2,4-Trimethylbenzene	2.827	2.821	0.2	87	0.00
85 T	sec-Butylbenzene	3.359	3.246	3.4	87	0.00
86 T	p-Isopropyltoluene	3.038	2.997	1.3	87	0.00
87 T	1,3-Dichlorobenzene	1.599	1.545	3.4	90	0.00
88 T	1,4-Dichlorobenzene	1.618	1.554	4.0	91	0.00
89 T	n-Butylbenzene	2.678	2.727	-1.8	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090419\
Data File : VU034213.D
Acq On : 03 Sep 2019 09:27
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Sep 04 04:40:51 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.525	0.447	14.9	80	0.00
91 T	1,2-Dichlorobenzene	1.532	1.486	3.0	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.256	0.211	17.6	75	0.00
93 T	1,2,4-Trichlorobenzene	0.937	1.029	-9.8	89	0.00
94 T	Hexachlorobutadiene	0.599	0.530	11.5	82	0.00
95 T	Naphthalene	2.584	2.581	0.1	79	0.00
96 T	1,2,3-Trichlorobenzene	0.981	0.982	-0.1	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6