


Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090420\
 Data File : VU040037.D
 Acq On : 04 Sep 2020 08:41
 Operator : SY/MD
 Sample : BFB28
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB28

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Mon Sep 07 02:10:55 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.1	41104	PASS
75	95	30	80	58.0	103368	PASS
95	95	100	100	100.0	178112	PASS
96	95	5	9	6.8	12042	PASS
173	174	0.00	2	0.5	549	PASS
174	95	50	120	66.9	119112	PASS
175	174	5	9	8.1	9666	PASS
176	174	95	101	95.5	113794	PASS
177	176	5	9	6.4	7244	PASS

m/z	Abundance
43.90	261.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

m/z	Abundance
36.00	2233.0
37.00	12348.0
38.00	10656.0
39.00	4254.0
40.00	345.0
44.00	1764.0
45.00	2119.0
45.90	285.0
47.00	2186.0
47.90	1297.0
49.00	10163.0
50.00	43016.0
51.00	13504.0
52.00	589.0
55.00	697.0
56.00	4042.0
57.00	6514.0
58.10	469.0
60.00	2186.0
61.00	10901.0
62.00	11334.0
63.10	9489.0
64.00	1047.0
65.10	872.0
67.20	619.0
68.00	21040.0
69.10	21656.0
70.00	1880.0
72.00	1277.0
73.00	9999.0
74.10	37632.0
75.10	107272.0
76.10	9840.0
77.10	1319.0
78.00	755.0
78.90	10363.0
79.90	2758.0
81.00	10451.0
82.00	2358.0
83.10	279.0
85.90	264.0
87.00	7963.0
88.00	6363.0
91.00	1111.0
92.10	5998.0
93.10	8976.0
94.10	24624.0
95.10	182464.0
96.10	12757.0
97.10	355.0
103.90	1012.0
104.90	297.0
105.90	1062.0
107.00	231.0
109.90	190.0
110.90	361.0
111.80	171.0
112.90	200.0
114.80	322.0
115.90	748.0
117.00	1701.0
117.90	831.0
119.00	1290.0
123.90	195.0
125.80	118.0
127.90	663.0
129.00	305.0
129.90	833.0
130.90	337.0
135.00	329.0
137.00	449.0
139.80	138.0
141.00	2671.0
142.10	358.0
143.00	2684.0
144.00	102.0
145.00	230.0
146.00	247.0
148.00	420.0
148.90	104.0
149.80	131.0
154.90	356.0
157.00	369.0
158.80	197.0
160.90	289.0
170.50	223.0
172.10	1071.0
173.10	1648.0
174.10	119040.0
175.10	10064.0
176.10	113752.0
177.10	7417.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

178.10

183.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

m/z	Abundance
36.00	2191.0
37.00	12363.0
38.00	10616.0
39.00	4169.0
40.00	381.0
43.00	102.0
44.00	1760.0
45.00	1955.0
45.90	235.0
47.10	1935.0
47.90	1472.0
49.00	9913.0
50.00	42416.0
51.00	12928.0
52.00	555.0
55.00	783.0
56.00	4130.0
57.00	6585.0
58.10	360.0
60.00	2251.0
61.00	10748.0
62.00	11216.0
63.10	9364.0
64.00	1058.0
65.10	736.0
67.20	625.0
68.00	20800.0
69.10	21424.0
70.10	1715.0
71.10	106.0
72.10	1307.0
73.00	9942.0
74.10	37632.0
75.10	106960.0
76.10	9791.0
77.10	1293.0
77.90	783.0
78.90	10070.0
79.90	2908.0
81.00	10542.0
82.00	2108.0
82.90	244.0
83.20	234.0
85.90	235.0
87.00	7682.0
88.00	6630.0
91.00	1194.0
92.10	5908.0
93.10	9179.0
94.10	25008.0
95.10	184064.0
96.10	12408.0
97.10	310.0
103.90	1044.0
104.90	339.0
105.90	1065.0
107.00	241.0
109.80	191.0
110.90	308.0
111.80	190.0
112.90	198.0
114.80	302.0
115.90	795.0
117.00	1724.0
117.90	842.0
119.00	1204.0
123.90	173.0
125.80	113.0
128.00	656.0
129.00	356.0
130.00	834.0
130.90	368.0
135.00	351.0
135.90	117.0
137.00	473.0
139.00	105.0
139.80	124.0
141.00	2575.0
142.10	324.0
143.00	2674.0
144.10	106.0
145.00	217.0
146.00	271.0
147.00	108.0
148.00	451.0
150.10	136.0
152.80	102.0
154.90	440.0
157.00	466.0
158.80	216.0
161.00	293.0
170.20	194.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

170.50	200.0
171.10	170.0
172.20	1035.0
174.10	123080.0
175.10	10044.0
176.10	117648.0
177.10	7525.0
178.10	166.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

m/z	Abundance
36.00	1918.0
37.00	10880.0
38.00	9622.0
39.00	3716.0
40.00	328.0
43.00	103.0
44.00	1628.0
45.00	1551.0
45.90	149.0
47.10	1823.0
47.90	1430.0
49.00	9048.0
50.00	37880.0
51.00	11091.0
52.00	535.0
52.90	112.0
55.00	751.0
56.00	3714.0
57.00	5879.0
58.10	234.0
60.00	2060.0
61.10	9595.0
62.00	9884.0
63.00	8200.0
64.00	895.0
65.20	656.0
67.10	519.0
68.00	18976.0
69.00	19376.0
70.10	1540.0
72.00	1063.0
73.10	8592.0
74.10	33864.0
75.10	95872.0
76.10	8563.0
77.00	1057.0
77.90	772.0
78.90	8796.0
79.90	2643.0
81.00	9729.0
82.00	1782.0
82.80	246.0
83.20	190.0
86.20	222.0
87.00	6451.0
88.00	6182.0
91.00	1080.0
92.10	5384.0
93.00	8229.0
94.10	22512.0
95.10	167808.0
96.10	10961.0
97.10	271.0
104.00	958.0
104.90	338.0
105.90	969.0
107.00	216.0
109.80	183.0
110.90	230.0
111.90	179.0
112.90	166.0
114.80	265.0
115.90	854.0
117.00	1610.0
117.90	742.0
119.00	1092.0
123.90	150.0
128.00	628.0
129.00	345.0
130.00	801.0
130.90	346.0
135.00	354.0
135.90	127.0
137.00	390.0
139.00	155.0
139.90	122.0
141.00	2318.0
142.10	263.0
143.00	2401.0
145.00	231.0
146.00	289.0
147.00	104.0
147.90	455.0
150.10	132.0
154.90	446.0
157.00	445.0
158.80	192.0
161.10	283.0
170.20	167.0
170.60	131.0
171.00	172.0
172.10	983.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28

174.10	115216.0
175.10	8891.0
176.10	109984.0
177.10	6791.0
178.10	162.0

Instrument :
MSVOA_U
ClientSampleId :
BFB28