

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090424\
 Data File : VU060603.D
 Acq On : 04 Sep 2024 16:40
 Operator : MD/SY
 Sample : P3809-03 50X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YE3E9

Quant Time: Sep 05 03:57:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 05 03:55:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	129048	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	137283	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	55908	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	33439	5.867	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	117.400%
7) Chloroethane-d5	1.911	69	33960	3.932	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.600%
11) 1,1-Dichloroethene-d2	2.563	65	14920	4.441	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.800%
20) 2-Butanone-d5	4.637	46	83539	42.621	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.240%
24) Chloroform-d	5.055	84	72779	4.479	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
26) 1,2-Dichloroethane-d4	5.695	65	37821	4.460	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.200%
32) Benzene-d6	5.721	84	137193	4.570	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	6.685	67	46780	4.830	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.600%
41) Toluene-d8	7.894	98	109979	3.905	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.200%
43) trans-1,3-Dichloroprop...	8.177	79	13216	3.813	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.200%
46) 2-Hexanone-d5	8.631	63	56731	34.187	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	68.380%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	34407	3.992	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	38629	4.686	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.800%
Target Compounds						Qvalue
16) Methylene chloride	3.039	84	5222	0.508	ug/L	# 65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090424\
Data File : VU060603.D
Acq On : 04 Sep 2024 16:40
Operator : MD/SY
Sample : P3809-03 50X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YE3E9

Quant Time: Sep 05 03:57:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 05 03:55:13 2024
Response via : Initial Calibration

