

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_U\DATA\VU090519\  
 Data File : VU034262.D  
 Acq On : 04 Sep 2019 16:54  
 Operator : JC/SP  
 Sample : K4547-33  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 T7-S-G2-(0)

Quant Time: Sep 05 03:40:04 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\82U082919W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Aug 29 04:11:12 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 4.96  | 168  | 419808   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.87  | 114  | 665302   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 9.07  | 117  | 684379   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.46 | 152  | 332144   | 50.00 | ug/l  | 0.00     |

|                             |        |     |          |       |         |      |
|-----------------------------|--------|-----|----------|-------|---------|------|
| System Monitoring Compounds |        |     |          |       |         |      |
| 33) 1,2-Dichloroethane-d4   | 5.29   | 65  | 260928   | 48.51 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 97.02%  |      |
| 35) Dibromofluoromethane    | 4.86   | 113 | 225638   | 48.90 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 97.80%  |      |
| 50) Toluene-d8              | 7.55   | 98  | 888563   | 51.43 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 102.86% |      |
| 62) 4-Bromofluorobenzene    | 10.29  | 95  | 306862   | 45.66 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 91.32%  |      |


| Target Compounds               | R.T.  | QIon | Response | Conc      | Units  | Qvalue |
|--------------------------------|-------|------|----------|-----------|--------|--------|
| 3) Chloromethane               | 1.32  | 50   | 359      | Below Cal |        | 91     |
| 5) Bromomethane                | 1.63  | 94   | 521      | Below Cal |        | 97     |
| 11) Tert butyl alcohol         | 2.82  | 59   | 4377     | 3.529     | ug/l # | 73     |
| 13) Acrolein                   | 2.17  | 56   | 112      | 0.215     | ug/l # | 46     |
| 16) Acetone                    | 2.31  | 43   | 9109     | 3.319     | ug/l   | 99     |
| 18) Methyl Acetate             | 2.61  | 43   | 3714     | 0.524     | ug/l # | 90     |
| 20) Methylene Chloride         | 2.69  | 84   | 7326     | 1.410     | ug/l   | 93     |
| 31) Cyclohexane                | 4.96  | 56   | 7268     | 0.905     | ug/l # | 1      |
| 36) 1,1-Dichloropropene        | 4.96  | 75   | 29541    | 4.694     | ug/l # | 48     |
| 38) Carbon Tetrachloride       | 4.96  | 117  | 38316    | 6.308     | ug/l # | 16     |
| 43) Isopropyl Acetate          | 5.53  | 43   | 78668    | 7.338     | ug/l   | 97     |
| 48) Methyl methacrylate        | 6.53  | 41   | 4314     | 0.834     | ug/l # | 14     |
| 49) 1,4-Dioxane                | 6.68  | 88   | 148      | 1.142     | ug/l # | 1      |
| 51) 4-Methyl-2-Pentanone       | 7.40  | 43   | 63230    | 8.540     | ug/l # | 46     |
| 54) cis-1,3-Dichloropropene    | 7.40  | 75   | 34867    | 4.354     | ug/l # | 60     |
| 56) Ethyl methacrylate         | 7.96  | 69   | 129      | 0.987     | ug/l # | 27     |
| 57) 1,3-Dichloropropane        | 8.40  | 76   | 6180     | 0.734     | ug/l # | 42     |
| 59) 2-Hexanone                 | 8.40  | 43   | 77843    | 13.927    | ug/l # | 53     |
| 76) 1,2,3-Trichloropropane     | 10.29 | 75   | 146478   | 19.567    | ug/l # | 37     |
| 81) trans-1,4-Dichloro-2-buten | 10.29 | 75   | 146478   | 65.887    | ug/l # | 100    |
| 95) Naphthalene                | 13.82 | 128  | 41       | 1.792     | ug/l # | 70     |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA\_U\DATA\VU090519\  
Data File : VU034262.D  
Acq On : 04 Sep 2019 16:54  
Operator : JC/SP  
Sample : K4547-33  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
T7-S-G2-(0)

Quant Time: Sep 05 03:40:04 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\82U082919W.M  
Quant Title : SW846 8260  
QLast Update : Thu Aug 29 04:11:12 2019  
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.96 min Scan# 1205

Delta R.T. 0.01 min

Lab File: VU034262.D

Acq: 04 Sep 2019 16:54

Instrument :

MSVOA\_U

ClientSampleId :


T7-S-G2-(0)

Tgt Ion:168 Resp: 419808

Ion Ratio Lower Upper

168 100

99 52.2 42.3 63.5



Abundance Ion 167.90 (167.60 to 168.60): VU034018.D (-1185) (-)

Ion 98.90 (98.60 to 99.60): VU034018.D (-1185) (-)

4.96

200000



150000

100000

50000

0

Time-->



#3

Chloromethane

Concen: Below Cal

RT: 1.32 min Scan# 73

Delta R.T. 0.00 min

Lab File: VU034262.D


Acq: 04 Sep 2019 16:54

Tgt Ion: 50 Resp: 359

Ion Ratio Lower Upper

50 100

52 37.6 26.2 39.4



Abundance Ion 49.90 (49.60 to 50.60): VU034018.D (-63) (-)

Ion 51.90 (51.60 to 52.60): VU034018.D (-63) (-)

1.32

400


300

200

100

0

Time-->





#5

Bromomethane

Concen: Below Cal

RT: 1.63 min Scan# 169

Delta R.T. 0.04 min

Lab File: VU034262.D

Acq: 04 Sep 2019 16:54

Instrument :

MSVOA\_U

ClientSampleId :


T7-S-G2-(0)

Tgt Ion: 94 Resp: 521

Ion Ratio Lower Upper




94 100

96 98.7 76.3 114.5



Abundance Ion 93.90 (93.60 to 94.60): VU0

Ion 95.90 (95.60 to 96.60): VU0



#11

Tert butyl alcohol

Concen: 3.529 ug/l

RT: 2.82 min Scan# 539

Delta R.T. 0.00 min

Lab File: VU034262.D


Acq: 04 Sep 2019 16:54

Tgt Ion: 59 Resp: 4377

Ion Ratio Lower Upper



59 100


57 0.0 8.0 12.0#



Abundance Ion 59.00 (58.70 to 59.70): VU0

Ion 57.00 (56.70 to 57.70): VU0





#13

Acrolein

Concen: 0.215 ug/l

RT: 2.17 min Scan# 335

Delta R.T. -0.01 min

Lab File: VU034262.D

Acq: 04 Sep 2019 16:54

Instrument :

MSVOA\_U

ClientSampleId :


T7-S-G2-(0)

Tgt Ion: 56 Resp: 112




Ion Ratio Lower Upper

56 100

55 116.1 57.1 85.7#



Abundance Ion 56.00 (55.70 to 56.70): VU034262.D (-246) (-)



#16

Acetone

Concen: 3.319 ug/l

RT: 2.31 min Scan# 380

Delta R.T. 0.00 min

Lab File: VU034262.D


Acq: 04 Sep 2019 16:54

Tgt Ion: 43 Resp: 9109



Ion Ratio Lower Upper


43 100

58 34.1 26.8 40.2



Abundance Ion 43.00 (42.70 to 43.70): VU034262.D (-286) (-)









#18  
Methyl Acetate  
Concen: 0.524 ug/l  
RT: 2.61 min Scan# 473  
Delta R.T. 0.01 min  
Lab File: VU034262.D  
Acq: 04 Sep 2019 16:54




Instrument :  
MSVOA\_U  
ClientSampleId :  
T7-S-G2-(0)


Tgt Ion: 43 Resp: 3714  
Ion Ratio Lower Upper  
43 100  
74 19.0 19.2 28.8#



#20  
Methylene Chloride  
Concen: 1.410 ug/l  
RT: 2.69 min Scan# 497  
Delta R.T. 0.01 min  
Lab File: VU034262.D  
Acq: 04 Sep 2019 16:54

Tgt Ion: 84 Resp: 7326  
Ion Ratio Lower Upper  
84 100  
49 120.3 100.2 150.4  
51 30.9 30.8 46.2  
86 58.6 52.5 78.7





#31

Cyclohexane

Concen: 0.905 ug/l

RT: 4.96 min Scan# 1205

Delta R.T. -0.01 min

Lab File: VU034262.D

Acq: 04 Sep 2019 16:54

Instrument :  
MSVOA\_U  
ClientSampleId :  
T7-S-G2-(0)


Tgt Ion: 56 Resp: 7268

Ion Ratio Lower Upper




56 100

69 194.2 25.4 38.0#

84 176.0 65.8 98.8#



Abundance Ion 56.00 (55.70 to 56.70): VU0  
Ion 69.00 (68.70 to 69.70): VU0  
Ion 84.00 (83.70 to 84.70): VU0



#33

1,2-Dichloroethane-d4

Concen: 48.507 ug/l

RT: 5.29 min Scan# 1306

Delta R.T. 0.00 min

Lab File: VU034262.D


Acq: 04 Sep 2019 16:54

Tgt Ion: 65 Resp: 260928



Ion Ratio Lower Upper

65 100

67 55.2 0.0 110.6



Abundance Ion 64.90 (64.60 to 65.60): VU0  
Ion 66.90 (66.60 to 67.60): VU0









#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 5.87 min Scan# 1486  
 Delta R.T. 0.00 min  
 Lab File: VU034262.D  
 Acq: 04 Sep 2019 16:54

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 T7-S-G2-(0)

Tgt Ion:114 Resp: 665302  
 Ion Ratio Lower Upper  
 114 100  
 63 20.2 0.0 42.0  
 88 16.2 0.0 32.0




Abundance Ion 113.90 (113.60 to 114.60): V  
 Ion 63.00 (62.70 to 63.70): VU0  
 Ion 87.90 (87.60 to 88.60): VU0






#35  
 Dibromofluoromethane  
 Concen: 48.898 ug/l  
 RT: 4.86 min Scan# 1173  
 Delta R.T. 0.00 min  
 Lab File: VU034262.D  
 Acq: 04 Sep 2019 16:54

Tgt Ion:113 Resp: 225638  
 Ion Ratio Lower Upper  
 113 100  
 111 100.6 81.7 122.5  
 192 21.5 17.1 25.7



Abundance Ion 112.80 (112.50 to 113.50): V  
 Ion 110.80 (110.50 to 111.50): V  
 Ion 191.70 (191.40 to 192.40): V









#36  
 1,1-Dichloropropene  
 Concen: 4.694 ug/l  
 RT: 4.96 min Scan# 1204  
 Delta R.T. -0.15 min  
 Lab File: VU034262.D  
 Acq: 04 Sep 2019 16:54




Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 T7-S-G2-(0)


Tgt Ion: 75 Resp: 29541  
 Ion Ratio Lower Upper  
 75 100  
 110 7.6 17.8 53.5#  
 77 0.0 25.0 37.4#



#38  
 Carbon Tetrachloride  
 Concen: 6.308 ug/l  
 RT: 4.96 min Scan# 1204  
 Delta R.T. -0.15 min  
 Lab File: VU034262.D  
 Acq: 04 Sep 2019 16:54

Tgt Ion: 117 Resp: 38316  
 Ion Ratio Lower Upper  
 117 100  
 119 5.7 77.0 115.4#  
 121 0.0 23.9 35.9#









#43  
Isopropyl Acetate  
Concen: 7.338 ug/l  
RT: 5.53 min Scan# 1382  
Delta R.T. 0.01 min  
Lab File: VU034262.D  
Acq: 04 Sep 2019 16:54




Instrument :  
MSVOA\_U  
ClientSampleId :  
T7-S-G2-(0)


Tgt Ion: 43 Resp: 78668  
Ion Ratio Lower Upper  
43 100  
61 20.1 17.2 25.8  
87 12.6 10.8 16.2



#48  
Methyl methacrylate  
Concen: 0.834 ug/l  
RT: 6.53 min Scan# 1692  
Delta R.T. -0.07 min  
Lab File: VU034262.D  
Acq: 04 Sep 2019 16:54

Tgt Ion: 41 Resp: 4314  
Ion Ratio Lower Upper  
41 100  
69 0.0 69.5 104.3#  
39 0.0 40.3 60.5#





#49

1,4-Dioxane

Concen: 1.142 ug/l

RT: 6.68 min Scan# 1740

Delta R.T. 0.09 min

Lab File: VU034262.D

Acq: 04 Sep 2019 16:54

Instrument :

MSVOA\_U

ClientSampleId :

T7-S-G2-(0)


Tgt Ion: 88 Resp: 148

Ion Ratio Lower Upper

88 100

43 984896.6 24.8 37.2#




58 5526.4 59.3 88.9#



Abundance Ion 88.00 (87.70 to 88.70): VU034262.D (-1740) (-)

Ion 43.00 (42.70 to 43.70): VU034262.D (-1740) (-)

Ion 58.00 (57.70 to 58.70): VU034262.D (-1740) (-)



#50

Toluene-d8

Concen: 51.426 ug/l

RT: 7.55 min Scan# 2008

Delta R.T. -0.00 min

Lab File: VU034262.D


Acq: 04 Sep 2019 16:54

Tgt Ion: 98 Resp: 888563

Ion Ratio Lower Upper



98 100


100 64.8 50.3 75.5



Abundance Ion 98.00 (97.70 to 98.70): VU034262.D (-2008) (-)

Ion 100.00 (99.70 to 100.70): VU034262.D (-2008) (-)





#51

4-Methyl-2-Pentanone

Concen: 8.540 ug/l

RT: 7.40 min Scan# 1963

Delta R.T. -0.04 min

Lab File: VU034262.D

Acq: 04 Sep 2019 16:54

Instrument :

MSVOA\_U

ClientSampleId :


T7-S-G2-(0)

Tgt Ion: 43 Resp: 63230

Ion Ratio Lower Upper

43 100

58 7.1 32.2 48.4#






Abundance Ion 43.00 (42.70 to 43.70): VU034018.D (-1956) (-)

Ion 58.00 (57.70 to 58.70): VU034018.D (-1956) (-)

7.40

Time-->



#54

cis-1,3-Dichloropropene

Concen: 4.354 ug/l

RT: 7.40 min Scan# 1963

Delta R.T. 0.15 min

Lab File: VU034262.D

Acq: 04 Sep 2019 16:54


Tgt Ion: 75 Resp: 34867

Ion Ratio Lower Upper

75 100

77 0.8 25.2 37.8#

39 24.4 34.6 51.8#





Abundance Ion 74.90 (74.60 to 75.60): VU034018.D (-1901) (-)


Ion 76.90 (76.60 to 77.60): VU034018.D (-1901) (-)

Ion 39.00 (38.70 to 39.70): VU034018.D (-1901) (-)

7.40

Time-->





#56

Ethyl methacrylate

Concen: 0.987 ug/l

RT: 7.96 min Scan# 2137

Delta R.T. -0.04 min

Lab File: VU034262.D

Acq: 04 Sep 2019 16:54

Instrument :

MSVOA\_U

ClientSampleId :

T7-S-G2-(0)


Tgt Ion: 69 Resp: 129

Ion Ratio Lower Upper

69 100

41 0.0 50.7 76.1#




39 0.0 24.2 36.2#



Abundance Ion 69.00 (68.70 to 69.70): VU034018.D (-2136) (-)

Ion 41.00 (40.70 to 41.70): VU034018.D (-2136) (-)

Ion 39.00 (38.70 to 39.70): VU034018.D (-2136) (-)



#57

1,3-Dichloropropane

Concen: 0.734 ug/l

RT: 8.40 min Scan# 2273

Delta R.T. 0.18 min

Lab File: VU034262.D


Acq: 04 Sep 2019 16:54

Tgt Ion: 76 Resp: 6180

Ion Ratio Lower Upper



76 100


78 0.0 26.2 39.4#



Abundance Ion 75.90 (75.60 to 76.60): VU034018.D (-2204) (-)

Ion 77.90 (77.60 to 78.60): VU034018.D (-2204) (-)









#59  
2-Hexanone  
Concen: 13.927 ug/l  
RT: 8.40 min Scan# 2273  
Delta R.T. 0.05 min  
Lab File: VU034262.D  
Acq: 04 Sep 2019 16:54




Instrument :  
MSVOA\_U  
ClientSampled :  
T7-S-G2-(0)


Tgt Ion: 43 Resp: 77843  
Ion Ratio Lower Upper  
43 100  
58 22.2 28.4 85.3#



#62  
4-Bromofluorobenzene  
Concen: 45.657 ug/l  
RT: 10.29 min Scan# 2861  
Delta R.T. -0.00 min  
Lab File: VU034262.D  
Acq: 04 Sep 2019 16:54

Tgt Ion: 95 Resp: 306862  
Ion Ratio Lower Upper  
95 100  
174 88.0 0.0 168.8  
176 83.5 0.0 161.8









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 9.07 min Scan# 2482  
Delta R.T. -0.00 min  
Lab File: VU034262.D  
Acq: 04 Sep 2019 16:54




Instrument :  
MSVOA\_U  
ClientSampleId :  
T7-S-G2-(0)


Tgt Ion:117 Resp: 684379  
Ion Ratio Lower Upper  
117 100  
82 55.0 45.1 67.7  
119 32.0 25.4 38.0



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 11.46 min Scan# 3226  
Delta R.T. -0.00 min  
Lab File: VU034262.D  
Acq: 04 Sep 2019 16:54

Tgt Ion:152 Resp: 332144  
Ion Ratio Lower Upper  
152 100  
115 56.6 41.1 123.3  
150 157.9 0.0 348.8





#76

1,2,3-Trichloropropane

Concen: 19.567 ug/l

RT: 10.29 min Scan# 2861

Delta R.T. -0.19 min

Lab File: VU034262.D

Acq: 04 Sep 2019 16:54

Instrument :

MSVOA\_U

ClientSampleId :


T7-S-G2-(0)

Tgt Ion: 75 Resp: 146478

Ion Ratio Lower Upper




75 100

77 1.3 20.1 60.2#



Abundance Ion 74.90 (74.60 to 75.60): VU0

Ion 77.00 (76.70 to 77.70): VU0



#81

trans-1,4-Dichloro-2-butene

Concen: 65.887 ug/l

RT: 10.29 min Scan# 2861

Delta R.T. -0.21 min

Lab File: VU034262.D

Acq: 04 Sep 2019 16:54


Tgt Ion: 75 Resp: 146478

Ion Ratio Lower Upper

75 100

53 0.2 0.0 0.0#



89 0.0 0.0 0.0




Abundance Ion 74.90 (74.60 to 75.60): VU0

Ion 53.00 (52.70 to 53.70): VU0

Ion 88.90 (88.60 to 89.60): VU0





#95

Naphthalene

Concen: 1.792 ug/l

RT: 13.82 min Scan# 3960

Delta R.T. 0.10 min

Lab File: VU034262.D


Acq: 04 Sep 2019 16:54

Instrument :

MSVOA\_U

ClientSampleId :

T7-S-G2-(0)




Tgt Ion:128 Resp: 41

Ion Ratio Lower Upper

128 100

127 0.0 10.2 15.4#

129 0.0 8.4 12.6#



Abundance

Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

