


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090519\
 Data File : VU034242.D
 Acq On : 04 Sep 2019 08:54
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Title : SW846 8260
 Last Update : Thu Aug 29 04:11:12 2019



AutoFind: Scans 2860, 2861, 2862; Background Corrected with Scan 2847

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	16155	PASS
75	95	30	60	47.0	43477	PASS
95	95	100	100	100.0	92586	PASS
96	95	5	9	6.6	6141	PASS
173	174	0.00	2	0.3	274	PASS
174	95	50	100	87.4	80890	PASS
175	174	5	9	7.3	5909	PASS
176	174	95	101	97.1	78568	PASS
177	176	5	9	6.1	4793	PASS