

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090523\
 Data File : VU055091.D
 Acq On : 05 Sep 2023 14:00
 Operator : MD/SY
 Sample : 04267-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH4F1

Quant Time: Sep 06 02:17:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 06 02:14:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	164159	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	169836	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	88773	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	36122	4.695	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	93.800%	
7) Chloroethane-d5	1.907	69	44504	4.997	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	100.000%	
11) 1,1-Dichloroethene-d2	2.560	65	22218	5.068	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	101.400%	
20) 2-Butanone-d5	4.624	46	139047	62.278	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	124.560%	
24) Chloroform-d	5.058	84	111058	5.440	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	108.800%	
26) 1,2-Dichloroethane-d4	5.701	65	57271	5.607	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	112.200%	
32) Benzene-d6	5.724	84	214603	5.119	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.400%	
36) 1,2-Dichloropropane-d6	6.688	67	71813	5.386	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	107.800%	
41) Toluene-d8	7.897	98	191036	4.947	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.000%	
43) trans-1,3-Dichloroprop...	8.180	79	18868	5.138	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	102.800%	
46) 2-Hexanone-d5	8.634	63	86791	58.220	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	116.440%	
56) 1,1,2,2-Tetrachloroeth...	10.756	84	54014	5.760	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	115.200%	
66) 1,2-Dichlorobenzene-d4	12.193	152	70553	5.445	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	108.800%	
Target Compounds					Qvalue	
13) Acetone	2.628	43	76478	51.108	ug/L	98
15) Methyl Acetate	2.952	43	27146	8.863	ug/L	98
16) Methylene chloride	3.042	84	6555	0.565	ug/L	95
21) 2-Butanone	4.724	43	12984	5.906	ug/L	93
40) 4-Methyl-2-pentanone	7.791	43	10731	1.912	ug/L #	89
42) Toluene	7.971	91	25019	0.649	ug/L	95
52) Ethylbenzene	9.579	91	5760	0.134	ug/L	89
53) m,p-Xylene	9.695	106	5106	0.320	ug/L	85
54) o-Xylene	10.103	106	3080	0.196	ug/L	100
55) Styrene	10.119	104	13536	0.522	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090523\
Data File : VU055091.D
Acq On : 05 Sep 2023 14:00
Operator : MD/SY
Sample : 04267-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH4F1

Quant Time: Sep 06 02:17:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 06 02:14:32 2023
Response via : Initial Calibration

