

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090524\
 Data File : VU060625.D
 Acq On : 05 Sep 2024 15:10
 Operator : MD/SY
 Sample : P3842-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0UE7

Quant Time: Sep 06 02:06:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 06 02:03:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	120512	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	127777	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	52427	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	27930	5.247	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.000%
7) Chloroethane-d5	1.911	69	30216	3.747	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.000%
11) 1,1-Dichloroethene-d2	2.560	65	13190	4.205	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.000%
20) 2-Butanone-d5	4.644	46	82656	45.158	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.320%
24) Chloroform-d	5.055	84	66684	4.394	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%
26) 1,2-Dichloroethane-d4	5.698	65	33731	4.260	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
32) Benzene-d6	5.721	84	124958	4.472	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
36) 1,2-Dichloropropane-d6	6.685	67	42792	4.747	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.000%
41) Toluene-d8	7.894	98	100685	3.841	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.800%
43) trans-1,3-Dichloroprop...	8.177	79	12158	3.769	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.400%
46) 2-Hexanone-d5	8.631	63	54384	35.210	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.420%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	32726	4.080	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	36880	4.771	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.400%
Target Compounds						
16) Methylene chloride	3.039	84	123210	12.843	ug/L	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090524\
Data File : VU060625.D
Acq On : 05 Sep 2024 15:10
Operator : MD/SY
Sample : P3842-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0UE7

Quant Time: Sep 06 02:06:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 06 02:03:06 2024
Response via : Initial Calibration

