

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090524\
 Data File : VU060616.D
 Acq On : 05 Sep 2024 11:36
 Operator : MD/SY
 Sample : P3809-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YE3F0

Quant Time: Sep 06 02:04:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 06 02:03:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	125003	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	133919	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	51707	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	32994	5.976	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	119.600%
7) Chloroethane-d5	1.911	69	32379	3.871	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.400%
11) 1,1-Dichloroethene-d2	2.560	65	14364	4.414	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.200%
20) 2-Butanone-d5	4.637	46	87011	45.829	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.660%
24) Chloroform-d	5.058	84	72801	4.625	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
26) 1,2-Dichloroethane-d4	5.695	65	37702	4.590	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.800%
32) Benzene-d6	5.721	84	134098	4.579	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.685	67	46264	4.897	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.000%
41) Toluene-d8	7.891	98	110526	4.023	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.400%
43) trans-1,3-Dichloroprop...	8.180	79	13430	3.972	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.400%
46) 2-Hexanone-d5	8.630	63	57911	35.774	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	71.540%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	35132	4.179	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	38798	5.089	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090524\
Data File : VU060616.D
Acq On : 05 Sep 2024 11:36
Operator : MD/SY
Sample : P3809-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YE3F0

Quant Time: Sep 06 02:04:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 06 02:03:06 2024
Response via : Initial Calibration

