

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090618\
 Data File : VU026607.D
 Acq On : 07 Sep 2018 03:01
 Operator : MD/SY
 Sample : J4765-01 50X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AE3

Quant Time: Sep 07 07:28:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 07 05:26:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	152767	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	144419	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	59162	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	52513	55.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	110.96%
7) Chloroethane-d5	1.68	69	47645	55.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	111.48%
11) 1,1-Dichloroethene-d2	2.27	63	77034	36.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.28%
21) 2-Butanone-d5	4.18	46	86686	118.00	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	118.00%
24) Chloroform-d	4.65	84	94771	53.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.02%
26) 1,2-Dichloroethane-d4	5.31	65	64461	54.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.32%
32) Benzene-d6	5.34	84	191875	57.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	115.28%
36) 1,2-Dichloropropane-d6	6.33	67	66665	58.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	116.60%
41) Toluene-d8	7.57	98	167765	52.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.42%
43) trans-1,3-Dichloropropene-	7.85	79	26584	50.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.82%
47) 2-Hexanone-d5	8.31	63	61012	116.34	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	116.34%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	90437	53.68	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.36%
64) 1,2-Dichlorobenzene-d4	11.86	152	64231	59.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	119.68%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090618\
Data File : VU026607.D
Acq On : 07 Sep 2018 03:01
Operator : MD/SY
Sample : J4765-01 50X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AE3

Quant Time: Sep 07 07:28:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 07 05:26:58 2018
Response via : Initial Calibration

