

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090618\
 Data File : VU026574.D
 Acq On : 06 Sep 2018 12:35
 Operator : MD/SY
 Sample : VU0906WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK46

Quant Time: Sep 07 04:50:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 07 04:44:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	163910	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	151722	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	64205	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	53734	52.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.82%
7) Chloroethane-d5	1.68	69	45808	49.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.88%
11) 1,1-Dichloroethene-d2	2.27	63	76525	33.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.92%
21) 2-Butanone-d5	4.18	46	77962	98.91	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.91%
24) Chloroform-d	4.65	84	88233	46.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.00%
26) 1,2-Dichloroethane-d4	5.31	65	60931	47.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.42%
32) Benzene-d6	5.34	84	182607	52.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.44%
36) 1,2-Dichloropropane-d6	6.33	67	61059	50.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.66%
41) Toluene-d8	7.57	98	160784	47.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.26%
43) trans-1,3-Dichloropropene-	7.85	79	26823	48.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.78%
47) 2-Hexanone-d5	8.31	63	54754	99.38	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.38%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	82526	46.63	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.26%
64) 1,2-Dichlorobenzene-d4	11.86	152	60451	51.89	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.78%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090618\
Data File : VU026574.D
Acq On : 06 Sep 2018 12:35
Operator : MD/SY
Sample : VU0906WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK46

Quant Time: Sep 07 04:50:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 07 04:44:56 2018
Response via : Initial Calibration

