

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090618\
 Data File : VU026576.D
 Acq On : 06 Sep 2018 13:24
 Operator : MD/SY
 Sample : J4694-21
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Sep 11 12:23:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 07 04:44:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	153000	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	143169	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	61510	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	49728	52.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.92%
7) Chloroethane-d5	1.68	69	43255	50.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.04%
11) 1,1-Dichloroethene-d2	2.27	63	73049	34.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.44%
21) 2-Butanone-d5	4.18	46	79867	108.55	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.55%
24) Chloroform-d	4.65	84	88646	49.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.02%
26) 1,2-Dichloroethane-d4	5.31	65	59796	50.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.32%
32) Benzene-d6	5.34	84	173065	52.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.90%
36) 1,2-Dichloropropane-d6	6.33	67	61608	54.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.70%
41) Toluene-d8	7.57	98	150968	47.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.78%
43) trans-1,3-Dichloropropene-	7.85	79	26501	51.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.38%
47) 2-Hexanone-d5	8.31	63	56852	109.35	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.35%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	84063	50.33	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.66%
64) 1,2-Dichlorobenzene-d4	11.86	152	60237	53.98	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.96%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090618\
Data File : VU026576.D
Acq On : 06 Sep 2018 13:24
Operator : MD/SY
Sample : J4694-21
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Sep 11 12:23:42 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 07 04:44:56 2018
Response via : Initial Calibration

