

Data File : VU034298.D
Acq On : 05 Sep 2019 20:57
Operator : JC/SP
Sample : K4548-14
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T8-S-G2-(0)

Quant Time: Sep 06 07:42:17 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	435366	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	687828	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	672670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	323968	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	274197	41.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.26%	
35) Dibromofluoromethane	4.86	113	243768	38.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.18%	
50) Toluene-d8	7.55	98	953433	39.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.30%	
62) 4-Bromofluorobenzene	10.29	95	332546	35.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	71.64%	


Target Compounds				Qvalue		
5) Bromomethane	1.64	94	712	Below Cal		87
6) Chloroethane	1.64	64	325	Below Cal	#	43
8) Diethyl Ether	2.09	74	1633	0.488	ug/l	94
10) Methyl Iodide	2.41	142	318	3.910	ug/l #	50
11) Tert butyl alcohol	2.83	59	4194	3.014	ug/l #	73
13) Acrolein	2.19	56	241	0.206	ug/l #	24
16) Acetone	2.31	43	11071	3.248	ug/l	95
18) Methyl Acetate	2.61	43	3882	0.468	ug/l	91
20) Methylene Chloride	2.69	84	11206	1.620	ug/l	95
31) Cyclohexane	4.96	56	7371	0.955	ug/l #	1
36) 1,1-Dichloropropene	4.96	75	29050	3.894	ug/l #	49
38) Carbon Tetrachloride	4.96	117	39070	4.905	ug/l #	15
43) Isopropyl Acetate	5.53	43	90586	6.584	ug/l	99
48) Methyl methacrylate	6.69	41	140747	21.538	ug/l #	37
49) 1,4-Dioxane	6.68	88	230	1.612	ug/l #	1
51) 4-Methyl-2-Pentanone	7.40	43	82228	8.634	ug/l #	45
54) cis-1,3-Dichloropropene	7.40	75	41809	4.042	ug/l #	63
56) Ethyl methacrylate	7.96	69	206	3.095	ug/l #	28
57) 1,3-Dichloropropane	8.40	76	7400	0.663	ug/l #	43
59) 2-Hexanone	8.40	43	91693	12.888	ug/l #	53
74) N-amyl acetate	9.92	43	1854	0.218	ug/l #	42
76) 1,2,3-Trichloropropane	10.29	75	157193	22.229	ug/l #	36
81) trans-1,4-Dichloro-2-buten	10.29	75	157193	74.753	ug/l #	100
93) 1,2,4-Trichlorobenzene	13.54	180	20	3.454	ug/l #	6
95) Naphthalene	13.79	128	292	3.978	ug/l #	69
96) 1,2,3-Trichlorobenzene	14.01	180	41	2.644	ug/l #	82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU034298.D
Acq On : 05 Sep 2019 20:57
Operator : JC/SP
Sample : K4548-14
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T8-S-G2-(0)

Quant Time: Sep 06 07:42:17 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.96 min Scan# 1204

Delta R.T. 0.00 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Instrument :

MSVOA_U

ClientSampleId :


T8-S-G2-(0)

Tgt Ion:168 Resp: 435366

Ion Ratio Lower Upper

168 100

99 52.2 41.6 62.4



Abundance Ion 167.90 (167.60 to 168.60): VU034274.D (-1186) (-)

Ion 98.90 (98.60 to 99.60): VU034274.D (-1186) (-)

4.96

200000

150000



100000

50000

0

4.90 5.00 5.10

Time-->



#5

Bromomethane

Concen: Below Cal

RT: 1.64 min Scan# 170

Delta R.T. 0.03 min

Lab File: VU034298.D


Acq: 05 Sep 2019 20:57

Tgt Ion: 94 Resp: 712

Ion Ratio Lower Upper

94 100

96 82.4 75.7 113.5



Abundance Ion 93.90 (93.60 to 94.60): VU034298.D (-68) (-)

Ion 95.90 (95.60 to 96.60): VU034298.D (-68) (-)

1.64

500

400

300


200


100

0

1.60 1.65

Time-->





#6

Chloroethane

Concen: Below Cal

RT: 1.64 min Scan# 170

Delta R.T. -0.05 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Instrument :

MSVOA_U

ClientSampleId :


T8-S-G2-(0)

Tgt Ion: 64 Resp: 325

Ion Ratio Lower Upper




64 100

66 0.0 25.0 37.6#



Abundance Ion 63.90 (63.60 to 64.60): VU0

Ion 65.90 (65.60 to 66.60): VU0



#8

Diethyl Ether

Concen: 0.488 ug/l

RT: 2.09 min Scan# 312

Delta R.T. 0.00 min

Lab File: VU034298.D


Acq: 05 Sep 2019 20:57

Tgt Ion: 74 Resp: 1633

Ion Ratio Lower Upper



74 100


45 102.1 48.2 144.6



Abundance Ion 74.00 (73.70 to 74.70): VU0

Ion 45.00 (44.70 to 45.70): VU0





#10

Methyl Iodide

Concen: 3.910 ug/l

RT: 2.41 min Scan# 411

Delta R.T. 0.01 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Instrument :

MSVOA_U

ClientSampleId :

T8-S-G2-(0)


Tgt Ion: 142 Resp: 318

Ion Ratio Lower Upper

142 100

127 6.0 32.1 48.1#




141 0.0 11.5 17.3#



Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V



#11

Tert butyl alcohol

Concen: 3.014 ug/l

RT: 2.83 min Scan# 540

Delta R.T. 0.01 min

Lab File: VU034298.D


Acq: 05 Sep 2019 20:57

Tgt Ion: 59 Resp: 4194

Ion Ratio Lower Upper



59 100


57 0.0 7.9 11.9#



Abundance Ion 59.00 (58.70 to 59.70): VU0

Ion 57.00 (56.70 to 57.70): VU0





#13

Acrolein

Concen: 0.206 ug/l

RT: 2.19 min Scan# 342

Delta R.T. 0.01 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Instrument :

MSVOA_U

ClientSampleId :





T8-S-G2-(0)

Tgt Ion: 56 Resp: 241

Ion Ratio Lower Upper

56 100

55 134.9 57.4 86.2#



#16

Acetone

Concen: 3.248 ug/l

RT: 2.31 min Scan# 380

Delta R.T. 0.00 min

Lab File: VU034298.D




Acq: 05 Sep 2019 20:57


Tgt Ion: 43 Resp: 11071

Ion Ratio Lower Upper

43 100

58 31.4 27.6 41.4









#18
Methyl Acetate
Concen: 0.468 ug/l
RT: 2.61 min Scan# 473
Delta R.T. 0.01 min
Lab File: VU034298.D
Acq: 05 Sep 2019 20:57




Instrument :
MSVOA_U
ClientSampleId :
T8-S-G2-(0)


Tgt Ion: 43 Resp: 3882
Ion Ratio Lower Upper
43 100
74 19.2 18.9 28.3



#20
Methylene Chloride
Concen: 1.620 ug/l
RT: 2.69 min Scan# 497
Delta R.T. 0.00 min
Lab File: VU034298.D
Acq: 05 Sep 2019 20:57

Tgt Ion: 84 Resp: 11206
Ion Ratio Lower Upper
84 100
49 125.3 96.0 144.0
51 42.7 30.0 45.0
86 61.6 51.4 77.0





#31

Cyclohexane

Concen: 0.955 ug/l

RT: 4.96 min Scan# 1205

Delta R.T. -0.01 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Instrument :

MSVOA_U

ClientSampleId :

T8-S-G2-(0)


Tgt Ion: 56 Resp: 7371

Ion Ratio Lower Upper

56 100

69 201.9 26.7 40.1#




84 164.9 68.2 102.4#



Abundance Ion 56.00 (55.70 to 56.70): VU034298.D (-1188) (-)

Ion 69.00 (68.70 to 69.70): VU034298.D (-1188) (-)

Ion 84.00 (83.70 to 84.70): VU034298.D (-1188) (-)



#33

1,2-Dichloroethane-d4

Concen: 41.626 ug/l

RT: 5.29 min Scan# 1306

Delta R.T. 0.00 min

Lab File: VU034298.D


Acq: 05 Sep 2019 20:57

Tgt Ion: 65 Resp: 274197

Ion Ratio Lower Upper



65 100


67 56.8 0.0 112.4



Abundance Ion 64.90 (64.60 to 65.60): VU034298.D (-1212) (-)

Ion 66.90 (66.60 to 67.60): VU034298.D (-1212) (-)









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.87 min Scan# 1486
 Delta R.T. 0.00 min
 Lab File: VU034298.D
 Acq: 05 Sep 2019 20:57




Instrument :
 MSVOA_U
 ClientSampleId :
 T8-S-G2-(0)


Tgt Ion:114 Resp: 687828
 Ion Ratio Lower Upper
 114 100
 63 19.5 0.0 42.4
 88 15.3 0.0 32.4



#35
 Dibromofluoromethane
 Concen: 38.087 ug/l
 RT: 4.86 min Scan# 1173
 Delta R.T. 0.00 min
 Lab File: VU034298.D
 Acq: 05 Sep 2019 20:57

Tgt Ion:113 Resp: 243768
 Ion Ratio Lower Upper
 113 100
 111 101.2 81.5 122.3
 192 21.6 17.8 26.8





#36

1,1-Dichloropropene

Concen: 3.894 ug/l

RT: 4.96 min Scan# 1204

Delta R.T. -0.15 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Instrument :

MSVOA_U

ClientSampleId :

T8-S-G2-(0)





Tgt Ion: 75 Resp: 29050

Ion Ratio Lower Upper

75 100

110 9.4 18.3 54.9#

77 0.0 24.9 37.3#



#38

Carbon Tetrachloride

Concen: 4.905 ug/l

RT: 4.96 min Scan# 1204

Delta R.T. -0.15 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57




Tgt Ion: 117 Resp: 39070


Ion Ratio Lower Upper

117 100

119 4.9 77.8 116.6#

121 0.0 25.2 37.8#





#43

Isopropyl Acetate

Concen: 6.584 ug/l

RT: 5.53 min Scan# 1382

Delta R.T. 0.01 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Instrument :

MSVOA_U

ClientSampleId :

T8-S-G2-(0)





Tgt Ion: 43 Resp: 90586

Ion Ratio Lower Upper

43 100

61 20.9 17.1 25.7

87 13.9 11.1 16.7



#48

Methyl methacrylate

Concen: 21.538 ug/l

RT: 6.69 min Scan# 1741

Delta R.T. 0.08 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57




Tgt Ion: 41 Resp: 140747


Ion Ratio Lower Upper

41 100

69 0.1 72.8 109.2#

39 47.9 38.6 58.0





#49

1,4-Dioxane

Concen: 1.612 ug/l

RT: 6.68 min Scan# 1740

Delta R.T. 0.09 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Instrument :

MSVOA_U

ClientSampleId :

T8-S-G2-(0)


Tgt Ion: 88 Resp: 230

Ion Ratio Lower Upper

88 100

43 702738.7 23.8 35.6#




58 3579.6 58.4 87.6#



Abundance Ion 88.00 (87.70 to 88.70): VU034298.D (-1740) (-)

Ion 43.00 (42.70 to 43.70): VU034298.D (-1740) (-)

Ion 58.00 (57.70 to 58.70): VU034298.D (-1740) (-)



#50

Toluene-d8

Concen: 39.651 ug/l

RT: 7.55 min Scan# 2008

Delta R.T. 0.00 min

Lab File: VU034298.D


Acq: 05 Sep 2019 20:57

Tgt Ion: 98 Resp: 953433

Ion Ratio Lower Upper



98 100


100 65.3 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VU034298.D (-2008) (-)

Ion 100.00 (99.70 to 100.70): VU034298.D (-2008) (-)









#51
 4-Methyl-2-Pentanone
 Concen: 8.634 ug/l
 RT: 7.40 min Scan# 1963
 Delta R.T. -0.04 min
 Lab File: VU034298.D
 Acq: 05 Sep 2019 20:57




Instrument :
 MSVOA_U
 ClientSampleId :
 T8-S-G2-(0)


Tgt Ion: 43 Resp: 82228
 Ion Ratio Lower Upper
 43 100
 58 6.8 33.0 49.4#



#54
 cis-1,3-Dichloropropene
 Concen: 4.042 ug/l
 RT: 7.40 min Scan# 1963
 Delta R.T. 0.15 min
 Lab File: VU034298.D
 Acq: 05 Sep 2019 20:57

Tgt Ion: 75 Resp: 41809
 Ion Ratio Lower Upper
 75 100
 77 0.4 25.1 37.7#
 39 25.3 31.2 46.8#





#56

Ethyl methacrylate

Concen: 3.095 ug/l

RT: 7.96 min Scan# 2137

Delta R.T. -0.04 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57





Instrument :

MSVOA_U

ClientSampleId :

T8-S-G2-(0)

Tgt Ion:	69.00	Resp:	206
Ion	Ratio	Lower	Upper
69	100		
41	0.0	49.5	74.3#
39	0.0	22.5	33.7#



#57

1,3-Dichloropropane

Concen: 0.663 ug/l




RT: 8.40 min Scan# 2273


Delta R.T. 0.18 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Tgt Ion:	76	Resp:	7400
Ion	Ratio	Lower	Upper
76	100		
78	0.0	25.4	38.2#









#59
2-Hexanone
Concen: 12.888 ug/l
RT: 8.40 min Scan# 2273
Delta R.T. 0.05 min
Lab File: VU034298.D
Acq: 05 Sep 2019 20:57




Instrument :
MSVOA_U
ClientSampleId :
T8-S-G2-(0)


Tgt Ion: 43 Resp: 91693
Ion Ratio Lower Upper
43 100
58 22.6 28.8 86.5#



#62
4-Bromofluorobenzene
Concen: 35.819 ug/l
RT: 10.29 min Scan# 2861
Delta R.T. 0.00 min
Lab File: VU034298.D
Acq: 05 Sep 2019 20:57

Tgt Ion: 95 Resp: 332546
Ion Ratio Lower Upper
95 100
174 88.8 0.0 175.2
176 83.2 0.0 168.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.07 min Scan# 2482
Delta R.T. 0.00 min
Lab File: VU034298.D
Acq: 05 Sep 2019 20:57

Instrument :
MSVOA_U
ClientSampleId :
T8-S-G2-(0)

Tgt Ion:117 Resp: 672670
Ion Ratio Lower Upper
117 100
82 55.2 44.6 67.0
119 32.2 25.9 38.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VU0
Ion 118.90 (118.60 to 119.60): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.46 min Scan# 3226
Delta R.T. 0.00 min
Lab File: VU034298.D
Acq: 05 Sep 2019 20:57

Tgt Ion:152 Resp: 323968
Ion Ratio Lower Upper
152 100
115 56.5 39.1 117.3
150 158.6 0.0 345.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#74

N-ethyl acetate

Concen: 0.218 ug/l

RT: 9.92 min Scan# 2746

Delta R.T. -0.07 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Instrument :

MSVOA_U

ClientSampleId :

T8-S-G2-(0)

Tgt Ion: 43 Resp: 1854


Ion Ratio Lower Upper

43 100

70 0.0 36.4 54.6#

55 3.7 21.2 31.8#

61 0.0 20.4 30.6#






Abundance Ion 43.00 (42.70 to 43.70): VU034274.D (-2753) (-)

Ion 70.00 (69.70 to 70.70): VU034274.D (-2753) (-)

Ion 55.00 (54.70 to 55.70): VU034274.D (-2753) (-)

Ion 61.00 (60.70 to 61.70): VU034274.D (-2753) (-)



#76

1,2,3-Trichloropropane

Concen: 22.229 ug/l

RT: 10.29 min Scan# 2861

Delta R.T. -0.19 min

Lab File: VU034298.D


Acq: 05 Sep 2019 20:57

Tgt Ion: 75 Resp: 157193

Ion Ratio Lower Upper



75 100


77 1.3 20.8 62.2#



Abundance Ion 74.90 (74.60 to 75.60): VU034274.D (-2909) (-)

Ion 77.00 (76.70 to 77.70): VU034274.D (-2909) (-)





#81

trans-1,4-Dichloro-2-butene

Concen: 74.753 ug/l

RT: 10.29 min Scan# 2861

Delta R.T. -0.20 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57

Instrument :

MSVOA_U

ClientSampleId :

T8-S-G2-(0)





Tgt Ion: 75 Resp: 157193

Ion Ratio Lower Upper

75 100

53 0.1 0.0 0.0#

89 0.0 0.0 0.0



#93

1,2,4-Trichlorobenzene

Concen: 3.454 ug/l

RT: 13.54 min Scan# 3872

Delta R.T. 0.05 min

Lab File: VU034298.D

Acq: 05 Sep 2019 20:57




Tgt Ion: 180 Resp: 20


Ion Ratio Lower Upper

180 100

182 200.0 47.9 143.8#

145 0.0 14.4 43.0#









#95
Naphthalene
Concen: 3.978 ug/l
RT: 13.79 min Scan# 3951
Delta R.T. 0.07 min
Lab File: VU034298.D
Acq: 05 Sep 2019 20:57

Instrument :
MSVOA_U
ClientSampleId :
T8-S-G2-(0)

Tgt Ion:128 Resp: 292
Ion Ratio Lower Upper
128 100
127 0.0 10.6 15.8#
129 0.0 8.4 12.6#



#96
1,2,3-Trichlorobenzene
Concen: 2.644 ug/l
RT: 14.01 min Scan# 4019
Delta R.T. 0.05 min
Lab File: VU034298.D
Acq: 05 Sep 2019 20:57

Tgt Ion:180 Resp: 41
Ion Ratio Lower Upper
180 100
182 102.4 48.4 145.0
145 0.0 15.4 46.4#

