

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090619\
 Data File : VU034294.D
 Acq On : 05 Sep 2019 19:24
 Operator : JC/SP
 Sample : K4548-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T11-2-(0)

Quant Time: Sep 06 08:26:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	437254	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	694710	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	683495	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	340126	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	276523	41.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.60%	
35) Dibromofluoromethane	4.86	113	240956	37.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.54%	
50) Toluene-d8	7.55	98	971637	40.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.02%	
62) 4-Bromofluorobenzene	10.29	95	344367	36.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.46%	
Target Compounds						
16) Acetone	2.31	43	11592	3.386	ug/l	99
43) Isopropyl Acetate	5.53	43	84652	6.092	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090619\
Data File : VU034294.D
Acq On : 05 Sep 2019 19:24
Operator : JC/SP
Sample : K4548-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T11-2-(0)

Quant Time: Sep 06 08:26:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

