

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090619\
 Data File : VU034298.D
 Acq On : 05 Sep 2019 20:57
 Operator : JC/SP
 Sample : K4548-14
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T8-S-G2-(0)

Quant Time: Sep 06 08:32:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	435366	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	687828	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	672670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	323968	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	274197	41.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.26%	
35) Dibromofluoromethane	4.86	113	243768	38.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.18%	
50) Toluene-d8	7.55	98	953433	39.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.30%	
62) 4-Bromofluorobenzene	10.29	95	332546	35.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.64%	
Target Compounds						
16) Acetone	2.31	43	11071	3.248	ug/l	95
20) Methylene Chloride	2.69	84	11206	1.620	ug/l	95
43) Isopropyl Acetate	5.53	43	90586	6.584	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090619\
Data File : VU034298.D
Acq On : 05 Sep 2019 20:57
Operator : JC/SP
Sample : K4548-14
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T8-S-G2-(0)

Quant Time: Sep 06 08:32:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

