

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090619\
 Data File : VU034300.D
 Acq On : 05 Sep 2019 21:44
 Operator : JC/SP
 Sample : K4548-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T8-S-G3-(0)

Quant Time: Sep 06 08:36:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	432749	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	665697	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	684849	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	338996	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.29	65	277440	42.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.74%	
35) Dibromofluoromethane	4.86	113	242566	39.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.32%	
50) Toluene-d8	7.55	98	966160	41.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.04%	
62) 4-Bromofluorobenzene	10.29	95	344306	38.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.32	43	11548	3.408	ug/l	99
20) Methylene Chloride	2.69	84	8661	1.170	ug/l	97
43) Isopropyl Acetate	5.53	43	89177	6.697	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090619\
Data File : VU034300.D
Acq On : 05 Sep 2019 21:44
Operator : JC/SP
Sample : K4548-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T8-S-G3-(0)

Quant Time: Sep 06 08:36:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

