

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090619\
 Data File : VU034310.D
 Acq On : 06 Sep 2019 01:37
 Operator : JC/SP
 Sample : K4579-28
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PAR-12-B3-(0)

Quant Time: Sep 06 08:55:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	421398	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	653454	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	667142	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	329948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	272978	42.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.62%	
35) Dibromofluoromethane	4.86	113	238342	39.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.40%	
50) Toluene-d8	7.55	98	945180	41.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.74%	
62) 4-Bromofluorobenzene	10.29	95	333402	37.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.60%	
Target Compounds						
16) Acetone	2.31	43	12705	3.851	ug/l	98
20) Methylene Chloride	2.69	84	11077	1.663	ug/l	94
43) Isopropyl Acetate	5.53	43	68421	5.235	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090619\
Data File : VU034310.D
Acq On : 06 Sep 2019 01:37
Operator : JC/SP
Sample : K4579-28
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PAR-12-B3-(0)

Quant Time: Sep 06 08:55:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

