

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090622\
 Data File : VU050644.D
 Acq On : 06 Sep 2022 19:18
 Operator : SY/MD
 Sample : N4482-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW364

Quant Time: Sep 07 01:42:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 07 01:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	207526	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	207543	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	80231	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	43441	24.474	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	48.940%#
7) Chloroethane-d5	1.903	69	40496	33.738	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	67.480%#
11) 1,1-Dichloroethene-d2	2.562	63	73654	23.482	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	46.960%#
21) 2-Butanone-d5	4.623	46	114931	88.501	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.500%
24) Chloroform-d	5.063	84	150450	44.722	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.440%
26) 1,2-Dichloroethane-d4	5.700	65	102911	50.061	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.120%
32) Benzene-d6	5.726	84	280318	40.871	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.740%
36) 1,2-Dichloropropane-d6	6.690	67	98445	43.718	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.440%
41) Toluene-d8	7.899	98	226614	38.637	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.280%#
43) trans-1,3-Dichloroprop...	8.179	79	41680	44.735	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.460%
47) 2-Hexanone-d5	8.635	63	46908	82.815	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.820%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	156441	45.909	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.820%
66) 1,2-Dichlorobenzene-d4	12.192	152	87885	46.793	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.580%
Target Compounds						
13) Acetone	2.629	43	7881	8.045	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090622\
Data File : VU050644.D
Acq On : 06 Sep 2022 19:18
Operator : SY/MD
Sample : N4482-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW364

Quant Time: Sep 07 01:42:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 07 01:37:50 2022
Response via : Initial Calibration

