

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090622\
 Data File : VU050633.D
 Acq On : 06 Sep 2022 14:52
 Operator : SY/MD
 Sample : N4483-08
 Misc : 4.57g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW356ME

Quant Time: Sep 07 01:39:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 07 01:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	396086	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	273001	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	95780	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	161810	47.763	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	95.520%	
7) Chloroethane-d5	1.903	69	141786	61.890	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	123.780%	
11) 1,1-Dichloroethene-d2	2.562	63	207683	34.691	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	69.380%	
21) 2-Butanone-d5	4.623	46	265290	107.033	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	107.030%	
24) Chloroform-d	5.063	84	327306	50.976	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.960%	
26) 1,2-Dichloroethane-d4	5.700	65	211560	53.920	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	107.840%	
32) Benzene-d6	5.726	84	652258	72.298	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	144.600%#	
36) 1,2-Dichloropropane-d6	6.690	67	225734	76.209	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	152.420%#	
41) Toluene-d8	7.899	98	536343	69.519	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	139.040%#	
43) trans-1,3-Dichloroprop...	8.179	79	86837	70.854	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	141.700%#	
47) 2-Hexanone-d5	8.632	63	104887	140.776	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	140.780%#	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	198365	44.254	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	88.500%	
66) 1,2-Dichlorobenzene-d4	12.192	152	115312	51.429	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.860%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090622\
Data File : VU050633.D
Acq On : 06 Sep 2022 14:52
Operator : SY/MD
Sample : N4483-08
Misc : 4.57g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW356ME

Quant Time: Sep 07 01:39:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 07 01:37:50 2022
Response via : Initial Calibration

