

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090622\
 Data File : VU050637.D
 Acq On : 06 Sep 2022 16:33
 Operator : SY/MD
 Sample : N4399-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 07 01:39:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 07 01:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	221838	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	218912	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	86434	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	60505	31.888	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	63.780%
7) Chloroethane-d5	1.903	69	51826	40.391	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.780%
11) 1,1-Dichloroethene-d2	2.562	63	104166	31.067	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.140%
21) 2-Butanone-d5	4.620	46	124302	89.542	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.540%
24) Chloroform-d	5.063	84	173808	48.332	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.660%
26) 1,2-Dichloroethane-d4	5.700	65	112854	51.356	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.720%
32) Benzene-d6	5.726	84	332569	45.971	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.940%
36) 1,2-Dichloropropane-d6	6.690	67	112873	47.522	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.040%
41) Toluene-d8	7.899	98	282771	45.708	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.420%
43) trans-1,3-Dichloroprop...	8.179	79	45780	46.584	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.160%
47) 2-Hexanone-d5	8.632	63	50656	84.788	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.790%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	172898	48.103	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	12.195	152	101249	50.040	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090622\
Data File : VU050637.D
Acq On : 06 Sep 2022 16:33
Operator : SY/MD
Sample : N4399-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 07 01:39:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 07 01:37:50 2022
Response via : Initial Calibration

