

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090623\
 Data File : VU055158.D
 Acq On : 06 Sep 2023 22:57
 Operator : MD/SY
 Sample : 04286-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Sep 07 02:27:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 06 05:03:37 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	169721	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.418	117	177765	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.814	152	90479	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	35402	4.451	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	89.000%	
7) Chloroethane-d5	1.907	69	44048	4.784	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.600%	
11) 1,1-Dichloroethene-d2	2.560	65	22276	4.915	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	98.200%	
20) 2-Butanone-d5	4.628	46	151305	65.547	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	131.100%#	
24) Chloroform-d	5.062	84	115004	5.448	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	109.000%	
26) 1,2-Dichloroethane-d4	5.701	65	60810	5.758	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	115.200%	
32) Benzene-d6	5.727	84	219982	5.013	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.200%	
36) 1,2-Dichloropropane-d6	6.689	67	75257	5.392	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	107.800%	
41) Toluene-d8	7.898	98	189221	4.681	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.600%	
43) trans-1,3-Dichloroprop...	8.180	79	17160	4.465	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.200%	
46) 2-Hexanone-d5	8.634	63	83986	53.825	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	107.660%	
56) 1,1,2,2-Tetrachloroeth...	10.756	84	57999	5.909	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	118.200%	
66) 1,2-Dichlorobenzene-d4	12.193	152	69699	5.278	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.600%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.515	50	3374	0.336	ug/L #	80
12) 1,1-Dichloroethene	2.567	96	1429	0.182	ug/L #	1
17) Methyl tert-butyl Ether	3.367	73	3092	0.150	ug/L #	93
19) 1,1-Dichloroethane	3.869	63	5272	0.306	ug/L #	93
22) cis-1,2-Dichloroethene	4.663	96	56107	5.758	ug/L	95
34) Trichloroethene	6.541	95	29427	2.799	ug/L	90
35) Methylcyclohexane	6.689	83	17578	1.191	ug/L #	19
37) 1,2-Dichloropropane	6.689	63	8615	0.836	ug/L #	88
47) Tetrachloroethene	8.553	164	1484	0.204	ug/L #	73
48) 2-Hexanone	8.682	43	24019	5.219	ug/L #	87
49) Dibromochloromethane	8.557	129	1264	0.167	ug/L #	9
51) Chlorobenzene	9.447	112	4628	0.173	ug/L #	72
61) 1,2,3-Trichloropropane	11.811	75	12386	2.063	ug/L #	70


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090623\
Data File : VU055158.D
Acq On : 06 Sep 2023 22:57
Operator : MD/SY
Sample : 04286-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :


Quant Time: Sep 07 02:27:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 06 05:03:37 2023
Response via : Initial Calibration








#3
Chloromethane
Concen: 0.336 ug/L
RT: 1.515 min Scan# 70
Delta R.T. 0.000 min
Lab File: VU055158.D
Acq: 06 Sep 2023 22:57

Instrument :
MSVOA_U
ClientSampleId :







Tgt Ion: 50 Resp: 3374
Ion Ratio Lower Upper
50 100
52 43.3 22.5 41.7#



#12
1,1-Dichloroethene
Concen: 0.182 ug/L
RT: 2.567 min Scan# 397
Delta R.T. -0.006 min
Lab File: VU055158.D
Acq: 06 Sep 2023 22:57


Tgt Ion: 96 Resp: 1429
Ion Ratio Lower Upper
96 100
61 586.0 120.8 224.4#
63 3720.4 94.5 175.5#








#17
Methyl tert-butyl Ether
Concen: 0.150 ug/L
RT: 3.367 min Scan# 644
Delta R.T. 0.003 min
Lab File: VU055158.D
Acq: 06 Sep 2023 22:57

Instrument :
MSVOA_U
ClientSampleId :







Tgt Ion: 73 Resp: 3092
Ion Ratio Lower Upper
73 100
43 14.9 16.2 24.2#
57 22.7 17.4 26.2



#19
1,1-Dichloroethane
Concen: 0.306 ug/L
RT: 3.869 min Scan# 802
Delta R.T. 0.003 min
Lab File: VU055158.D
Acq: 06 Sep 2023 22:57


Tgt Ion: 63 Resp: 5272
Ion Ratio Lower Upper
63 100
65 29.0 22.2 41.2
83 10.0 10.6 19.8#








#22
cis-1,2-Dichloroethene
Concen: 5.758 ug/L
RT: 4.663 min Scan# 1048
Delta R.T. 0.000 min
Lab File: VU055158.D
Acq: 06 Sep 2023 22:57

Instrument :
MSVOA_U
ClientSampleId :







Tgt Ion: 96 Resp: 56107
Ion Ratio Lower Upper
96 100
61 130.1 97.2 180.4
98 64.7 46.1 85.7



#34
Trichloroethene
Concen: 2.799 ug/L
RT: 6.541 min Scan# 1633
Delta R.T. 0.000 min
Lab File: VU055158.D
Acq: 06 Sep 2023 22:57

Tgt Ion: 95 Resp: 29427
Ion Ratio Lower Upper
95 100
97 55.4 44.7 83.1
132 81.4 66.6 123.6
130 89.0 66.5 123.5





#35

Methylcyclohexane

Concen: 1.191 ug/L

RT: 6.689 min Scan# 1

Delta R.T. -0.071 min

Lab File: VU055158.D

Acq: 06 Sep 2023 22:57

Instrument :

MSVOA_U

ClientSampleId :





Tgt Ion: 83 Resp: 17578

Ion Ratio Lower Upper

83 100

55 0.2 60.4 90.6#

98 0.0 35.3 52.9#



#37

1,2-Dichloropropane

Concen: 0.836 ug/L

RT: 6.689 min Scan# 1679

Delta R.T. -0.100 min

Lab File: VU055158.D




Acq: 06 Sep 2023 22:57


Tgt Ion: 63 Resp: 8615

Ion Ratio Lower Upper

63 100


112 0.0 3.2 4.8#








#47
Tetrachloroethene
Concen: 0.204 ug/L
RT: 8.553 min Scan# 21
Delta R.T. 0.000 min
Lab File: VU055158.D
Acq: 06 Sep 2023 22:57

Instrument :
MSVOA_U
ClientSampleId :







Tgt Ion:164 Resp: 1484
Ion Ratio Lower Upper
164 100
129 71.1 69.2 128.6
131 67.2 68.8 127.8#
166 104.1 92.3 171.5



#48
2-Hexanone
Concen: 5.219 ug/L
RT: 8.682 min Scan# 2299
Delta R.T. 0.000 min
Lab File: VU055158.D
Acq: 06 Sep 2023 22:57

Tgt Ion: 43 Resp: 24019
Ion Ratio Lower Upper
43 100
58 44.1 44.0 66.0
57 16.2 16.2 24.2
100 8.3 9.0 13.4#





#49

Dibromochloromethane

Concen: 0.167 ug/L

RT: 8.557 min Scan# 21

Delta R.T. -0.251 min

Lab File: VU055158.D

Acq: 06 Sep 2023 22:57

Instrument :

MSVOA_U





ClientSampleId :

Tgt Ion:129 Resp: 1264

Ion Ratio Lower Upper

129 100

127 0.0 55.9 103.7#



#51

Chlorobenzene

Concen: 0.173 ug/L

RT: 9.447 min Scan# 2537

Delta R.T. 0.000 min

Lab File: VU055158.D

Acq: 06 Sep 2023 22:57




Tgt Ion:112 Resp: 4628


Ion Ratio Lower Upper

112 100

114 33.3 21.7 40.3


77 91.8 48.6 72.8#





#61
1,2,3-Trichloropropane
Concen: 2.063 ug/L
RT: 11.811 min Scan# 31
Delta R.T. 0.990 min
Lab File: VU055158.D
Acq: 06 Sep 2023 22:57

Instrument :
MSVOA_U
ClientSampleId :



Tgt Ion: 75 Resp: 12386
Ion Ratio Lower Upper
75 100
110 1.6 28.2 42.2#
77 32.1 26.2 39.4

