

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090623\
 Data File : VU055155.D
 Acq On : 06 Sep 2023 21:44
 Operator : MD/SY
 Sample : 04286-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH4F4

Quant Time: Sep 06 22:30:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 06 05:03:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	170986	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	175510	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.814	152	90229	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	33037	4.123	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.400%
7) Chloroethane-d5	1.907	69	42701	4.603	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.000%
11) 1,1-Dichloroethene-d2	2.560	65	20797	4.555	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.000%
20) 2-Butanone-d5	4.628	46	136362	58.637	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.280%
24) Chloroform-d	5.062	84	105043	4.940	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.800%
26) 1,2-Dichloroethane-d4	5.702	65	54813	5.152	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.000%
32) Benzene-d6	5.724	84	202893	4.683	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
36) 1,2-Dichloropropane-d6	6.689	67	66929	4.857	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.200%
41) Toluene-d8	7.898	98	175077	4.387	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.800%
43) trans-1,3-Dichloroprop...	8.181	79	16119	4.248	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.000%
46) 2-Hexanone-d5	8.634	63	75535	49.031	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.060%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	51331	5.297	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.000%
66) 1,2-Dichlorobenzene-d4	12.193	152	63865	4.849	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.000%
Target Compounds						
3) Chloromethane	1.515	50	2592	0.256	ug/L	92
9) Trichlorofluoromethane	2.136	101	4203	0.271	ug/L	89
12) 1,1-Dichloroethene	2.573	96	57311	7.250	ug/L	89
17) Methyl tert-butyl Ether	3.361	73	14669	0.708	ug/L #	94
18) trans-1,2-Dichloroethene	3.354	96	1145	0.143	ug/L	81
19) 1,1-Dichloroethane	3.869	63	8698	0.502	ug/L	96
22) cis-1,2-Dichloroethene	4.663	96	91557	9.326	ug/L	96
29) 1,1,1-Trichloroethane	5.316	97	2425	0.160	ug/L #	84
34) Trichloroethene	6.537	95	494649	47.648	ug/L	97
47) Tetrachloroethene	8.553	164	27031	3.767	ug/L	97
51) Chlorobenzene	9.451	112	4140	0.157	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090623\
Data File : VU055155.D
Acq On : 06 Sep 2023 21:44
Operator : MD/SY
Sample : 04286-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH4F4

Quant Time: Sep 06 22:30:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 06 05:03:37 2023
Response via : Initial Calibration

