

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090623\
 Data File : VU055163.D
 Acq On : 07 Sep 2023 00:58
 Operator : MD/SY
 Sample : 04286-17
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH4G2

Quant Time: Sep 07 02:28:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 06 05:03:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	166914	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	171060	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	88844	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	29414	3.760	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.200%
7) Chloroethane-d5	1.907	69	37575	4.149	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.000%
11) 1,1-Dichloroethene-d2	2.560	65	19468	4.368	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.400%
20) 2-Butanone-d5	4.628	46	124094	54.663	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.320%
24) Chloroform-d	5.058	84	103631	4.992	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
26) 1,2-Dichloroethane-d4	5.701	65	54253	5.224	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.724	84	194159	4.598	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
36) 1,2-Dichloropropane-d6	6.689	67	67248	5.007	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.200%
41) Toluene-d8	7.898	98	166328	4.276	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.600%
43) trans-1,3-Dichloroprop...	8.180	79	15478	4.185	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.800%
46) 2-Hexanone-d5	8.634	63	71567	47.664	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.320%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	49690	5.261	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.200%
66) 1,2-Dichlorobenzene-d4	12.193	152	62329	4.806	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.200%
Target Compounds						
					Qvalue	
3) Chloromethane	1.518	50	1937	0.196	ug/L	99
5) Vinyl chloride	1.599	62	5407	0.519	ug/L	95
22) cis-1,2-Dichloroethene	4.666	96	65668	6.852	ug/L	93
34) Trichloroethene	6.541	95	208577	20.614	ug/L	96
47) Tetrachloroethene	8.553	164	1750	0.250	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090623\
Data File : VU055163.D
Acq On : 07 Sep 2023 00:58
Operator : MD/SY
Sample : 04286-17
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH4G2

Quant Time: Sep 07 02:28:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 06 05:03:37 2023
Response via : Initial Calibration

