

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090623\
 Data File : VU055138.D
 Acq On : 06 Sep 2023 14:19
 Operator : MD/SY
 Sample : 04267-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH4C9

Quant Time: Sep 06 20:41:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 06 05:03:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	216155	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	218485	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.814	152	113027	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	32994	3.257	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.200%
7) Chloroethane-d5	1.907	69	40675	3.469	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	69.400%
11) 1,1-Dichloroethene-d2	2.560	65	20862	3.614	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.200%
20) 2-Butanone-d5	4.624	46	127670	43.427	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.860%
24) Chloroform-d	5.062	84	104415	3.884	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.600%
26) 1,2-Dichloroethane-d4	5.702	65	55220	4.106	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.200%
32) Benzene-d6	5.724	84	198990	3.690	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.800%
36) 1,2-Dichloropropane-d6	6.689	67	68335	3.984	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.600%
41) Toluene-d8	7.898	98	173895	3.500	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.000%
43) trans-1,3-Dichloroprop...	8.181	79	17495	3.704	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.000%
46) 2-Hexanone-d5	8.634	63	78099	40.724	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.440%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	52216	4.328	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.600%
66) 1,2-Dichlorobenzene-d4	12.193	152	64303	3.898	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	78.000%#
Target Compounds						Qvalue
3) Chloromethane	1.518	50	2071	0.162	ug/L	89
17) Methyl tert-butyl Ether	3.364	73	5737	0.219	ug/L #	81
22) cis-1,2-Dichloroethene	4.666	96	54685	4.406	ug/L	99
25) Chloroform	5.087	83	15159	0.664	ug/L	89
34) Trichloroethene	6.541	95	41910	3.243	ug/L	98
47) Tetrachloroethene	8.553	164	3354	0.376	ug/L	88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090623\
Data File : VU055138.D
Acq On : 06 Sep 2023 14:19
Operator : MD/SY
Sample : 04267-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH4C9

Quant Time: Sep 06 20:41:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 06 05:03:37 2023
Response via : Initial Calibration









#3
Chloromethane
Concen: 0.162 ug/L
RT: 1.518 min Scan# 70
Delta R.T. 0.003 min
Lab File: VU055138.D
Acq: 06 Sep 2023 14:19




Instrument :
MSVOA_U
ClientSampleId :
BH4C9


Tgt Ion: 50 Resp: 2071
Ion Ratio Lower Upper
50 100
52 38.4 22.5 41.7



#17
Methyl tert-butyl Ether
Concen: 0.219 ug/L
RT: 3.364 min Scan# 645
Delta R.T. 0.000 min
Lab File: VU055138.D
Acq: 06 Sep 2023 14:19

Tgt Ion: 73 Resp: 5737
Ion Ratio Lower Upper
73 100
43 14.2 16.2 24.2#
57 9.9 17.4 26.2#









#22
 cis-1,2-Dichloroethene
 Concen: 4.406 ug/L
 RT: 4.666 min Scan# 1049
 Delta R.T. 0.003 min
 Lab File: VU055138.D
 Acq: 06 Sep 2023 14:19




Instrument :
 MSVOA_U
 ClientSampleId :
 BH4C9


Tgt Ion:	96	Resp:	54685
Ion	Ratio	Lower	Upper
96	100		
61	138.1	97.2	180.4
98	64.8	46.1	85.7



#25
 Chloroform
 Concen: 0.664 ug/L
 RT: 5.087 min Scan# 1181
 Delta R.T. 0.000 min
 Lab File: VU055138.D
 Acq: 06 Sep 2023 14:19

Tgt Ion:	83	Resp:	15159
Ion	Ratio	Lower	Upper
83	100		
85	58.5	47.2	87.6









#34
Trichloroethene
Concen: 3.243 ug/L
RT: 6.541 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU055138.D
Acq: 06 Sep 2023 14:19

Instrument :
MSVOA_U
ClientSampleId :
BH4C9

Tgt Ion: 95 Resp: 41910

Ion	Ratio	Lower	Upper
95	100		
97	64.5	44.7	83.1
132	91.7	66.6	123.6
130	92.6	66.5	123.5



#47
Tetrachloroethene
Concen: 0.376 ug/L
RT: 8.553 min Scan# 2259
Delta R.T. 0.000 min
Lab File: VU055138.D
Acq: 06 Sep 2023 14:19

Tgt Ion: 164 Resp: 3354

Ion	Ratio	Lower	Upper
164	100		
129	106.2	69.2	128.6
131	110.5	68.8	127.8
166	112.9	92.3	171.5

