

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090718\
 Data File : VU026611.D
 Acq On : 07 Sep 2018 09:33
 Operator : MD/SY
 Sample : VU0907WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0907WBL01

Quant Time: Sep 07 12:19:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U082018DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Mon Aug 20 13:35:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.75	96	24240	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	6559	0.75	ug/l	0.00
Spiked Amount 1.000			Recovery	=	75.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	7909	0.78	ug/l	0.00
Spiked Amount 1.000			Recovery	=	78.00%	

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090718\
Data File : VU026611.D
Acq On : 07 Sep 2018 09:33
Operator : MD/SY
Sample : VU0907WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0907WBL01

Quant Time: Sep 07 12:19:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U082018DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Mon Aug 20 13:35:41 2018
Response via : Initial Calibration

