

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090719\
 Data File : VU034327.D
 Acq On : 06 Sep 2019 13:52
 Operator : JC/SP
 Sample : K4722-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GW-200-B4-13.0-18.0-090519

Quant Time: Sep 06 14:15:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	336680	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	502954	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	488667	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	307215	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	253252	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
35) Dibromofluoromethane	4.86	113	222626	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
50) Toluene-d8	7.55	98	862552	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
62) 4-Bromofluorobenzene	10.29	95	312547	46.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.08%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluoroet	2.28	101	8044	2.064	ug/l	99
16) Acetone	2.31	43	19343	7.338	ug/l	97
30) Chloroform	4.65	83	183600	22.147	ug/l	100
44) Trichloroethene	6.17	130	143201	29.402	ug/l	97
47) Bromodichloromethane	6.74	83	17811	2.679	ug/l	99
64) Tetrachloroethene	8.21	164	2119	0.500	ug/l	85
95) Naphthalene	13.73	128	120697	9.393	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090719\
Data File : VU034327.D
Acq On : 06 Sep 2019 13:52
Operator : JC/SP
Sample : K4722-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GW-200-B4-13.0-18.0-090519

Quant Time: Sep 06 14:15:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

