

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090719\
 Data File : VU034328.D
 Acq On : 06 Sep 2019 14:15
 Operator : JC/SP
 Sample : K4722-15
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GW-200-B4-47.0-52.0-090519

Quant Time: Sep 06 16:06:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	316336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	466461	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	455542	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	280051	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.29	65	233068	48.70	ug/l	0.00
Spiked Amount			Recovery	=	97.40%	
35) Dibromofluoromethane	4.86	113	203751	46.94	ug/l	0.00
Spiked Amount			Recovery	=	93.88%	
50) Toluene-d8	7.55	98	771651	47.32	ug/l	0.00
Spiked Amount			Recovery	=	94.64%	
62) 4-Bromofluorobenzene	10.29	95	278552	44.24	ug/l	0.00
Spiked Amount			Recovery	=	88.48%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.31	43	4723	1.907	ug/l	97
21) trans-1,2-Dichloroethene	2.97	96	48366	13.335	ug/l	97
27) cis-1,2-Dichloroethene	4.20	96	191395	43.449	ug/l	87
30) Chloroform	4.66	83	13843	1.777	ug/l	96
40) Benzene	5.38	78	9162	0.561	ug/l #	89
44) Trichloroethene	6.17	130	822675	182.126	ug/l	99
68) m/p-Xylenes	9.37	106	1429	0.201	ug/l	77
69) o-Xylene	9.76	106	4261	0.612	ug/l	98
95) Naphthalene	13.74	128	53773	6.618	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090719\
Data File : VU034328.D
Acq On : 06 Sep 2019 14:15
Operator : JC/SP
Sample : K4722-15
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GW-200-B4-47.0-52.0-090519

Quant Time: Sep 06 16:06:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

