

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090719\
 Data File : VU034332.D
 Acq On : 06 Sep 2019 15:48
 Operator : JC/SP
 Sample : K4722-15DL 5X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GW-200-B4-47.0-52.0-090519DL

Quant Time: Sep 06 17:25:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	294811	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	435096	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	425650	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	259142	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	213841	47.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.88%	
35) Dibromofluoromethane	4.86	113	187893	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
50) Toluene-d8	7.55	98	720426	47.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.72%	
62) 4-Bromofluorobenzene	10.29	95	258783	44.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.14%	
Target Compounds						
21) trans-1,2-Dichloroethene	2.97	96	8400	2.485	ug/l	94
27) cis-1,2-Dichloroethene	4.21	96	34723	8.458	ug/l	85
30) Chloroform	4.66	83	2956	0.407	ug/l	89
44) Trichloroethene	6.17	130	145444	34.520	ug/l	98
95) Naphthalene	13.74	128	45175	6.374	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090719\
Data File : VU034332.D
Acq On : 06 Sep 2019 15:48
Operator : JC/SP
Sample : K4722-15DL 5X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GW-200-B4-47.0-52.0-090519DL

Quant Time: Sep 06 17:25:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

