

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090719\
 Data File : VU034335.D
 Acq On : 06 Sep 2019 16:57
 Operator : JC/SP
 Sample : K4663-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 OK-03-090519

Quant Time: Sep 06 17:37:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	311049	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	467936	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	469716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	284905	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	248837	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
35) Dibromofluoromethane	4.86	113	219850	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
50) Toluene-d8	7.55	98	836868	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	10.29	95	301726	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
Target Compounds						
16) Acetone	2.31	43	11903	4.887	ug/l	97
20) Methylene Chloride	2.69	84	25264	5.981	ug/l	97
43) Isopropyl Acetate	5.53	43	74345	7.943	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090719\
Data File : VU034335.D
Acq On : 06 Sep 2019 16:57
Operator : JC/SP
Sample : K4663-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OK-03-090519

Quant Time: Sep 06 17:37:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

