

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090719\
 Data File : VU034354.D
 Acq On : 07 Sep 2019 01:04
 Operator : JC/SP
 Sample : K4725-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S13-ER-2019-1

Quant Time: Sep 07 05:12:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	295160	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	429627	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	428661	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	254737	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	226147	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
35) Dibromofluoromethane	4.86	113	206246	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
50) Toluene-d8	7.55	98	760504	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
62) 4-Bromofluorobenzene	10.29	95	276090	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
Target Compounds						
					Qvalue	
16) Acetone	2.31	43	6011	2.601	ug/l	96
17) Carbon Disulfide	2.47	76	2537	0.352	ug/l #	81
20) Methylene Chloride	2.69	84	18831	4.612	ug/l	97
25) 2-Butanone	4.28	43	6461	2.094	ug/l	99
29) Tetrahydrofuran	4.62	42	236636	126.057	ug/l	99
89) n-Butylbenzene	11.87	91	4495	0.342	ug/l	90
95) Naphthalene	13.74	128	17962	4.940	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090719\
Data File : VU034354.D
Acq On : 07 Sep 2019 01:04
Operator : JC/SP
Sample : K4725-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S13-ER-2019-1

Quant Time: Sep 07 05:12:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

